

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-003072 Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 685 psi Regulator: 680 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: MP5290 Lot #: 21340 Exp. Date: 8/24/23
 - b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.300 AC
Sim. Ser #: MP3067 Lot #: 21420 Exp. Date: 10/20/23

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 1656 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 31 Exp. Date: 3/5/23

Test 1 <u>0.079</u>	Test 4 <u>0.080</u>	Test 7 <u>0.079</u>
Test 2 <u>0.078</u>	Test 5 <u>0.079</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.080</u>	Test 9 <u>0.079</u>
Average <u>0.079</u>		

C. Remarks/Maintenance: Annual inspection
Replaced breath tube + simulator return o-rings.
Replaced external battery.

Instrument is acceptable to be used in the field.

Janelle Postochiller
Breath Analyst Signature

6/24/22
Date

[Signature]
Reviewed by

062422
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.14.00 09/16
06/24/2022 11:58

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:59
02 Std. Sol.	0.018	12:00
03 Room Air	0.000	12:00
04 Std. Sol.	0.018	12:01
05 Room Air	0.000	12:02
06 Std. Sol.	0.018	12:02
07 Room Air	0.000	12:03


08 Sim Temp = 34.0°C

Simul Ser No = MP5290

Std Sol No = 21340

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.14.00 09/16
06/24/2022 12:05

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:05
02 Std. Sol.	0.295	12:06
03 Room Air	0.000	12:07
04 Std. Sol.	0.298	12:07
05 Room Air	0.000	12:08
06 Std. Sol.	0.297	12:09
07 Room Air	0.000	12:09

08 Sim Temp = 34.0°C

Simul Ser No = MP3067
Std Sol No = 21420
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High Ac check 0.300*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.14.00 09/16
06/24/2022 12:27

Test	AC	Time
01 Diagnostic	OK	12:28
02 Room Air	0.000	12:28
03 *Subject Test	RFI*	12:29
04 Room Air	0.000	12:29

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/TES989643
Lot No = 02621080A1
Cyl No = 31
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.14.00 09/16
06/24/2022 12:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:35
02 Std. Gas	0.080	12:35
03 Room Air	0.000	12:35
04 Std. Gas	0.079	12:36
05 Room Air	0.000	12:36
06 Std. Gas	0.080	12:37
07 Room Air	0.000	12:37

Lot No = 02621080A1
Cyl No = 31
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry calibration check # 4-6*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.14.00 09/16
06/24/2022 12:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:38
02 Std. Gas	0.079	12:38
03 Room Air	0.000	12:39
04 Std. Gas	0.079	12:39
05 Room Air	0.000	12:40
06 Std. Gas	0.079	12:40
07 Room Air	0.000	12:41

Lot No = 02621080A1
Cyl No = 31
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #7-9*