



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number <i>80-004199</i>	Instrument Location <i>Fargo Police PD3</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C). *ACA test was received 063022 jpr*
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): *NONE*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i> 198 118104	Date <i>6-24-22</i>
--	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Janelle Pettschiller</i>	Certified Date <i>06/30/2022</i>
--	-------------------------------------

Note: I called Troy Nielsen on 21 July 22 to change the location code on the instrument. The location code was changed from F6PD to FAPD. Troy verified no subject tests have been run on the instrument since the installation. 21 July 22 jpr

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = FGPD 8164.14.00 09/16
06/24/2022 *Location Code* 09:20
FAPD 21041777 JEL

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 13
Standard Expiration? 03/05/2023
Oper No? 110104

Flow Cal. Date: 06/17/2019
Slope 593
Intercept -481610

IR Calibration Date: 06/17/2019
3um 9um

0th Coef(*100): -17512 -19237
1st Coef(*100): 277207 136841
2nd Coef(*100): 1620 1027
H2O adj(mg/l*10k): 211 375

***** Printer Test End *****


Operator Signature

TROY NIELSEN

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = FGPD 8164.14.00 09/16
06/24/2022 09:24

*Location code
FAPD 21 July 22 gnt*

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:24
02 Std. Gas	0.077	09:25
03 Room Air	0.000	09:25
04 Std. Gas	0.078	09:26
05 Room Air	0.000	09:26
06 Std. Gas	0.077	09:27
07 Room Air	0.000	09:27

Lot No = 02621080A1

Cyl No = 13

Exp Date = 03/05/2023

County = 09

Oper No. = 110104


Operator Signature
TROY NIELSEN

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = FGPD 8164.14.00 09/16
06/24/2022 09:20

*Location code
FAPD 21 July 22 JTA*

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:21
02 Std. Gas	0.077	09:21
03 Room Air	0.000	09:22
04 Std. Gas	0.077	09:22
05 Room Air	0.000	09:23
06 Std. Gas	0.078	09:23
07 Room Air	RFI*	09:23
08 Room Air	RFI*	09:23

*Invalid Test
Inhibited - RFI

Lot No = 02621080A1
Cyl No = 13
Exp Date = 03/05/2023
County = 09 *RFI* Oper No. = 110104

Troy Nielsen
Operator Signature
TROY NIELSEN

Remarks: