

**INTOXILYZER® 8000 INSPECTION**

Intoxilyzer® 8000 Serial Number: 80-003059 Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 901 psi Regulator: 900 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
    - a.  Low AC (Level 1,C): Known Value  $\leq$  0.03 AC: 0.020 AC  
Sim. Ser #: MP6036 Lot #: 21340 Exp. Date: 8/24/23
    - b.  High AC (Level 1,C): Known Value  $\geq$  0.25 AC: 0.300 AC  
Sim. Ser #: MP3068 Lot #: 202201F Exp. Date: 1/18/24

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3847 Lot #: 1CS 6 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 34921080A1 Cyl #: 47 Exp. Date: 2/5/24

Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.080</u>	Test 5 <u>0.081</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.081</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

C. Remarks/Maintenance: Annual inspection.  
Replaced breath tube + simulator return o-rings.  
Replaced external battery.

Instrument is acceptable to be used in the field.

Jennifer Pertschiller  
Breath Analyst Signature  
[Signature]  
Reviewed by

06/27/2022  
Date  
06/29/22  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003059  
Location = TOXL                    8164.14.00 09/16  
06/27/2022                            10:48

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            24119080A1  
Standard Cyl #?            6  
Standard Expiration?       11/05/2021  
Oper No?                    132128

Flow Cal. Date:            11/28/2018  
Slope                       589  
Intercept                  -276335

IR Calibration Date:       12/04/2018  
                                  3um                            9um

-----  
0th Coef(\*100):        -25330                    -15825  
1st Coef(\*100):        278286                   137135  
2nd Coef(\*100):        1992                     1257  
H2O adj(mg/l\*10k):    525                      321

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = TOXL      8164.14.00 09/16  
06/27/2022      10:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:50
02 Std. Sol.	0.019	10:51
03 Room Air	0.000	10:52
04 Std. Sol.	0.019	10:52
05 Room Air	0.000	10:53
06 Std. Sol.	0.019	10:54
07 Room Air	0.000	10:54

08 Sim Temp = 34.0°C

Simul Ser No = MP6036  
Std Sol No = 21340  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = TOXL      8164.14.00 09/16  
06/27/2022      10:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:58
02 Std. Sol.	0.298	10:59
03 Room Air	0.000	10:59
04 Std. Sol.	0.299	11:00
05 Room Air	0.000	11:01
06 Std. Sol.	0.300	11:01
07 Room Air	0.000	11:02


08 Sim Temp = 34.0°C

Simul Ser No = MP3068

Std Sol No = 202201F

County = 08

Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *High AC check 0.300*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003059  
Location = TOXL                    8164.14.00 09/16  
06/27/2022                            11:41

Test	AC	Time
01 Room Air	0.000	11:42
02 *Subject Test	INT*	11:43
03 Room Air	0.000	11:44

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT CK  
Dr. Lic. = ND/DIS921456  
Lot No = 34921080A1  
Cyl No = 47  
Expiration Date = 02/05/2024  
County = 08                            Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks:

*Interferent Check*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003059  
Location = TOXL                    8164.14.00 09/16  
06/27/2022                            11:47

Test	AC	Time
01 Diagnostic	OK	11:48
02 Room Air	0.000	11:48
03 *Subject Test	RFI*	11:49
04 Room Air	0.000	11:50

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHK  
Dr. Lic. = ND/DIS921456  
Lot No = 34921080A1  
Cyl No = 47  
Expiration Date = 02/05/2024  
County = 08                            Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *RFI Check*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = TOXL      8164.14.00 09/16  
06/27/2022      11:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:51
02 Std. Gas	0.080	11:51
03 Room Air	0.000	11:51
04 Std. Gas	0.080	11:52
05 Room Air	0.000	11:52
06 Std. Gas	0.080	11:53
07 Room Air	0.000	11:53

Lot No = 34921080A1  
Cyl No = 47  
Exp Date = 02/05/2024  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003059  
Location = TOXL                    8164.14.00 09/16  
06/27/2022                            11:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:54
02 Std. Gas	0.080	11:54
03 Room Air	0.000	11:55
04 Std. Gas	0.081	11:55
05 Room Air	0.000	11:56
06 Std. Gas	0.081	11:56
07 Room Air	0.000	11:57

Lot No = 34921080A1

Cyl No = 47

Exp Date = 02/05/2024

County = 08

Oper No. = 132128

*Janelle Portscheller*

Operator Signature  
JANELLE PORTSCHELLER

Remarks:

*Dry Calibration. Check # 4-6*

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003059  
Location = TOXL                    8164.14.00 09/16  
06/27/2022                            11:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:58
02 Std. Gas	0.080	11:59
03 Room Air	0.000	11:59
04 Std. Gas	0.080	12:00
05 Room Air	0.000	12:00
06 Std. Gas	0.080	12:01
07 Room Air	0.000	12:01

Lot No = 34921080A1  
Cyl No = 47  
Exp Date = 02/05/2024  
County = 08                            Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 7-9*