

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004198

Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 492 psi Regulator: 560 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq$  0.03 AC: 0.020 AC  
Sim. Ser #: MP 5290 Lot #: 21340 Exp. Date: 8/24/23
- b.  High AC (Level 1,C): Known Value  $\geq$  0.25 AC: 0.300 AC  
Sim. Ser #: MP 3067 Lot #: 21420 Exp. Date: 10/20/23

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR 51149 Lot #: 16562 Exp. Date: NA

Display reads "Interferent Detect" 1657

3.  RFI Check (CMS Mode)  Display reads "RFI Detect"

4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC

Gas Cyl Lot #: 02621080A1 Cyl #: 31 Exp. Date: 3/5/23

Test 1 0.079 Test 4 0.078 Test 7 0.078

Test 2 0.078 Test 5 0.078 Test 8 0.077

Test 3 0.077 Test 6 0.077 Test 9 0.077

Average 0.078

DR 7345

C. Remarks/Maintenance: Annual inspection.

Replaced breath tube + simulator return o-rings.

Instrument is acceptable to be used in the field.

Janelle Pertschiller  
Breath Analyst Signature

06/28/2022  
Date

[Signature]  
Reviewed by

062522  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer           Alcohol Analyzer  
North Dakota Model 8000           SN 80-004198  
Location = TOXL                   8164.14.00 09/16  
06/28/2022                               11:21

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:           Enabled  
Start Test Sequence:       DABACABA  
Display Prelim Rslt?       Yes  
Display Third Digit?       Yes  
Inhib Printer(Y/N)?       No  
Display Volume?           No  
Disable On Memfull?       Yes  
# of Print Copies?         1  
Select Std (D/W/I)?       Dry  
Standard Value?           0.080  
Standard Lot #?           24119080A1  
Standard Cyl #?           33  
Standard Expiration?       11/05/2021  
Oper No?                   132128

Flow Cal. Date:           01/09/2012  
Slope                      637  
Intercept                 -467208

IR Calibration Date:      05/18/2016  
                              3um                   9um

	3um	9um
0th Coef(*100):	-19233	-23050
1st Coef(*100):	267031	132119
2nd Coef(*100):	2449	1142
H2O adj(mg/l*10k):	329	398

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004198  
Location = TOXL      8164.14.00 09/16  
06/28/2022      11:28

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:29
02 Std. Sol.	0.018	11:35
03 Room Air	0.000	11:35
04 Std. Sol.	0.019	11:36
05 Room Air	0.000	11:36
06 Std. Sol.	0.018	11:37
07 Room Air	0.000	11:38

08 Sim Temp = 34.0°C

Simul Ser No = MP5290

Std Sol No = 21340

County = 08

Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Low AC Check 0.020*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004198  
Location = TOXL                 8164.14.00 09/16  
06/28/2022                        11:39

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:39
02 Std. Sol.	0.295	11:41
03 Room Air	0.000	11:41
04 Std. Sol.	0.298	11:42
05 Room Air	0.000	11:42
06 Std. Sol.	0.298	11:43
07 Room Air	0.000	11:44

08 Sim Temp = 34.0°C

Simul Ser No = MP3067

Std Sol No = 21420

County = 08

Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *High Ac Check 0.300*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004198  
Location = TOXL                    8164.14.00 09/16  
06/28/2022                            11:51

Test	AC	Time
01 Room Air	0.000	11:52
02 *Subject Test	INT*	11:54
03 Room Air	0.000	11:54

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                      Weight = NA  
Test = DUI                            Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 02621080A1  
Cyl No = 31  
Expiration Date = 03/05/2023  
County = 08                          Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks:

*Interferent test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-004198  
Location = TOXL                              8164.14.00 09/16  
06/28/2022    12:05

Test	AC	Time
01 Diagnostic	OK	12:06
02 Room Air	0.000	12:07
03 *Subject Test	RFI*	12:07
04 Room Air	0.000	12:08

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                              Weight = NA  
Test = DUI                                      Cit = RFI CHK  
Dr. Lic. = ND/TES989643  
Lot No = 02621080A1  
Cyl No = 31  
Expiration Date = 03/05/2023  
County = 08                              Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *RFI check*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004198  
Location = TOXL                      8164.14.00 09/16  
06/28/2022                              12:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:09
02 Std. Gas	0.079	12:09
03 Room Air	0.000	12:10
04 Std. Gas	0.078	12:10
05 Room Air	0.000	12:11
06 Std. Gas	0.077	12:11
07 Room Air	0.000	12:12

Lot No = 02621080A1  
Cyl No = 31  
Exp Date = 03/05/2023  
County = 08                              Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004198  
Location = TOXL      8164.14.00 09/16  
06/28/2022      12:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:12
02 Std. Gas	0.078	12:13
03 Room Air	0.000	12:13
04 Std. Gas	0.078	12:13
05 Room Air	0.000	12:14
06 Std. Gas	0.077	12:14
07 Room Air	0.000	12:15

Lot No = 02621080A1  
Cyl No = 31  
Exp Date = 03/05/2023  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-6*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004198  
Location = TOXL                    8164.14.00 09/16  
06/28/2022                            12:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:16
02 Std. Gas	0.078	12:17
03 Room Air	0.000	12:17
04 Std. Gas	0.077	12:17
05 Room Air	0.000	12:18
06 Std. Gas	0.077	12:18
07 Room Air	0.000	12:19

Lot No = 02621080A1  
Cyl No = 31  
Exp Date = 03/05/2023  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 7-9*