

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 8U-002667

Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 856 psi Regulator: 850 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP6036 Lot #: 21340 Exp. Date: 8/24/23
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP3068 Lot #: 202201F Exp. Date: 1/18/24

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR 3847 Lot #: 1CS6 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 34921080A1 Cyl #: 47 Exp. Date: 2/15/24

Test 1	<u>0.080</u>	Test 4	<u>0.081</u>	Test 7	<u>0.080</u>
Test 2	<u>0.080</u>	Test 5	<u>0.080</u>	Test 8	<u>0.080</u>
Test 3	<u>0.081</u>	Test 6	<u>0.081</u>	Test 9	<u>0.080</u>
Average	<u>0.080</u>				

C. Remarks/Maintenance: Annual inspection.

Replaced breath tube and simulator return o-rings.

Instrument is acceptable to be used in the field.

Janelle Putschiller  
Breath Analyst Signature

06/28/2022  
Date

[Signature]  
Reviewed by

062922  
Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = TOXL      8164.14.00 09/16  
06/28/2022      11:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:21
02 Std. Sol.	0.020	11:30
03 Room Air	0.000	11:30
04 Std. Sol.	0.020	11:31
05 Room Air	0.000	11:31
06 Std. Sol.	0.021	11:32
07 Room Air	0.000	11:33

08 Sim Temp = 34.0°C

Simul Ser No = MP6036  
Std Sol No = 21340  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Low Ac check 0.020*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = TOXL      8164.14.00 09/16  
06/28/2022      11:35

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:36
02 Std. Sol.	0.300	11:37
03 Room Air	0.000	11:37
04 Std. Sol.	0.302	11:38
05 Room Air	0.000	11:39
06 Std. Sol.	0.302	11:39
07 Room Air	0.000	11:40

08 Sim Temp = 34.0°C

Simul Ser No = MP3068  
Std Sol No = 202201F  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *High Ac check 0.300*










Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = TOXL      8164.14.00 09/16  
06/28/2022      12:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:10
02 Std. Gas	0.081	12:11
03 Room Air	0.000	12:11
04 Std. Gas	0.080	12:12
05 Room Air	0.000	12:12
06 Std. Gas	0.081	12:12
07 Room Air	0.000	12:13

Lot No = 34921080A1  
Cyl No = 47  
Exp Date = 02/05/2024  
County = 08      Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry calibration check # 4-6*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = TOXL      8164.14.00 09/16  
06/28/2022      12:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:14
02 Std. Gas	0.080	12:14
03 Room Air	0.000	12:15
04 Std. Gas	0.080	12:15
05 Room Air	0.000	12:15
06 Std. Gas	0.080	12:16
07 Room Air	0.000	12:16

Lot No = 34921080A1  
Cyl No = 47  
Exp Date = 02/05/2024  
County = 08      Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check 7-9*