

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005941

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 957 psi Regulator: 950 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3070 Lot #: 22CP41184 Exp. Date: 3/1/24
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.400 AC
Sim. Ser #: MP 3065 Lot #: 207103E Exp. Date: 3/24/23

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: 1CS6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 34921080A1 Cyl #: 44 Exp. Date: 2/5/24

Test 1	<u>0.080</u>	Test 4	<u>0.080</u>	Test 7	<u>0.081</u>
Test 2	<u>0.079</u>	Test 5	<u>0.080</u>	Test 8	<u>0.081</u>
Test 3	<u>0.081</u>	Test 6	<u>0.080</u>	Test 9	<u>0.081</u>
Average	<u>0.080</u>				

C. Remarks/Maintenance: Annual inspection.

Replaced breath tube o-ring.
Per prior annual inspections, the instrument has an auto cal issue and is for classroom use only.

No Instrument is acceptable to be used in the field. No

Jennifer Putschiller
Breath Analyst Signature

06/28/22
Date

[Signature]
Reviewed by

062922
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.14.00 09/16
06/28/2022 10:33

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 24119080A1
Standard Cyl #? 5
Standard Expiration? 11/05/2021
Oper No? 132128

Flow Cal. Date: 05/30/2018
Slope 618
Intercept -570330

IR Calibration Date: 05/30/2018
 3um 9um

0th Coef(*100): -13785 -15807
1st Coef(*100): 281869 139821
2nd Coef(*100): 2733 1395
H2O adj(mg/l*10k): 308 362

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.14.00 09/16
06/28/2022 11:19

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:20
02 Std. Sol.	0.019	11:21
03 Room Air	0.000	11:22
04 Std. Sol.	0.019	11:22
05 Room Air	0.000	11:23
06 Std. Sol.	0.019	11:24
07 Room Air	0.000	11:24

08 Sim Temp = 34.0°C

Simul Ser No = MP3070
Std Sol No = 22CP41184
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.14.00 09/16
06/28/2022 11:34

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:35
02 Std. Sol.	0.405	11:36
03 Room Air	0.000	11:37
04 Std. Sol.	0.405	11:37
05 Room Air	0.000	11:38
06 Std. Sol.	0.405	11:39
07 Room Air	0.000	11:39

08 Sim Temp = 34.0°C

Simul Ser No = MP3065
Std Sol No = 202103E
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High AC check 0.400*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.14.00 09/16
06/28/2022 11:43

Test	AC	Time
01 Room Air	0.000	11:44
02 *Subject Test	INT*	11:45
03 Room Air	0.000	11:45

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CHK
Dr. Lic. = ND/DIS821456
Lot No = 34921080A1
Cyl No = 44
Expiration Date = 02/05/2024
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.14.00 09/16
06/28/2022 12:04

Test	AC	Time
01 Diagnostic	OK	12:05
02 Room Air	0.000	12:06
03 *Subject Test	RFI*	12:06
04 Room Air	0.000	12:06

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS821456
Lot No = 34921080A1
Cyl No = 44
Expiration Date = 02/05/2024
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI Check*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.14.00 09/16
06/28/2022 12:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:07
02 Std. Gas	0.080	12:08
03 Room Air	0.000	12:08
04 Std. Gas	0.079	12:09
05 Room Air	0.000	12:09
06 Std. Gas	0.081	12:09
07 Room Air	0.000	12:10

Lot No = 34921080A1
Cyl No = 44
Exp Date = 02/05/2024
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.14.00 09/16
06/28/2022 12:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:11
02 Std. Gas	0.080	12:11
03 Room Air	0.000	12:12
04 Std. Gas	0.080	12:12
05 Room Air	0.000	12:13
06 Std. Gas	0.080	12:13
07 Room Air	0.000	12:14

Lot No = 34921080A1
Cyl No = 44
Exp Date = 02/05/2024
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-6*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.14.00 09/16
06/28/2022 12:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:17
02 Std. Gas	0.081	12:17
03 Room Air	0.000	12:18
04 Std. Gas	0.081	12:18
05 Room Air	0.000	12:18
06 Std. Gas	0.081	12:19
07 Room Air	0.000	12:19

Lot No = 34921080A1
Cyl No = 44
Exp Date = 02/05/2024
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #7-9*