

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-003065 Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 977 psi Regulator: 950 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP6036 Lot #: 21340 Exp. Date: 5/24/23
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3068 Lot #: 202201F Exp. Date: 1/18/24

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR 3847 Lot #: 1CS 6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 34921080A1 Cyl #: 47 Exp. Date: 2/5/24

Test 1	<u>0.079</u>	Test 4	<u>0.079</u>	Test 7	<u>0.078</u>
Test 2	<u>0.077</u>	Test 5	<u>0.078</u>	Test 8	<u>0.077</u>
Test 3	<u>0.078</u>	Test 6	<u>0.079</u>	Test 9	<u>0.079</u>
Average	<u>0.078</u>				

C. Remarks/Maintenance: Annual inspection.
Replaced breath tube and simulator return o-rings.

Instrument is acceptable to be used in the field.

Janelle Portscheller
Breath Analyst Signature

6/28/22
Date

[Signature]
Reviewed by

062922
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.14.00 09/16
06/28/2022 15:05

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 26
Standard Expiration? 03/05/2023
Oper No? 132128

Flow Cal. Date: 05/19/2017
Slope 612
Intercept 21223

IR Calibration Date: 08/23/2016
 3um 9um

0th Coef(*100): -14934 -12309
1st Coef(*100): 289519 141483
2nd Coef(*100): 1834 736
H2O adj(mg/l*10k): 268 136

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.14.00 09/16
06/28/2022 15:08

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:09
02 Std. Sol.	0.020	15:10
03 Room Air	0.000	15:10
04 Std. Sol.	0.020	15:11
05 Room Air	0.000	15:11
06 Std. Sol.	0.021	15:12
07 Room Air	0.000	15:13

08 Sim Temp = 34.0°C

Simul Ser No = MP6036
Std Sol No = 21340
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.14.00 09/16
06/28/2022 15:16

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:16
02 Std. Sol.	0.298	15:17
03 Room Air	0.000	15:18
04 Std. Sol.	0.299	15:18
05 Room Air	0.000	15:19
06 Std. Sol.	0.298	15:20
07 Room Air	0.000	15:20

08 Sim Temp = 34.0°C

Simul Ser No = MP3068
Std Sol No = 202201
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High AC check 0.300*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.14.00 09/16
06/28/2022 16:19

Test	AC	Time
01 Diagnostic	OK	16:20
02 Room Air	0.000	16:21
03 *Subject Test	RFI*	16:22
04 Room Air	0.000	16:22

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS921456
Lot No = 34921080A1
Cyl No = 47
Expiration Date = 02/05/2024
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI Check*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.14.00 09/16
06/28/2022 16:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:29
02 Std. Gas	0.079	16:30
03 Room Air	0.000	16:30
04 Std. Gas	0.078	16:31
05 Room Air	0.000	16:31
06 Std. Gas	0.079	16:32
07 Room Air	0.000	16:32

Lot No = 34921080A1
Cyl No = 47
Exp Date = 02/05/2024
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-6*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.14.00 09/16
06/28/2022 16:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:35
02 Std. Gas	0.078	16:35
03 Room Air	0.000	16:36
04 Std. Gas	0.077	16:36
05 Room Air	0.000	16:37
06 Std. Gas	0.079	16:37
07 Room Air	0.000	16:38

Lot No = 34921080A1
Cyl No = 47
Exp Date = 02/05/2024
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #79*