

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-003069

Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 978 psi Regulator: 978 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
  - a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP 3070 Lot #: 22CP411P4 Exp. Date: 3/1/24
  - b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.400 AC  
Sim. Ser #: MP 3065 Lot #: 22A103E Exp. Date: 3/24/23

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR 3847 Lot #: 1CS6 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 34921080A1 Cyl #: 44 Exp. Date: 2/5/24

Test 1	<u>0.079</u>	Test 4	<u>0.079</u>	Test 7	<u>0.079</u>
Test 2	<u>0.079</u>	Test 5	<u>0.079</u>	Test 8	<u>0.079</u>
Test 3	<u>0.079</u>	Test 6	<u>0.079</u>	Test 9	<u>0.079</u>
Average <u>0.079</u>					

C. Remarks/Maintenance: Annual inspection  
Replaced breath tube and simulator return o-rings.  
Classroom only instrument. No battery status if left unplugged.

NO  Instrument is acceptable to be used in the field. NO

Janelle Pertschiller  
Breath Analyst Signature

06/28/22  
Date

Reviewed by \_\_\_\_\_

062922  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003069  
Location = TOXL                 8164.14.00 09/16  
06/28/2022                        15:06

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^ &\* ( ) \_ + ?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^ &\* ( ) \_ + ?

Current Instrument Setup

Data Entry Mode:                Enabled  
Start Test Sequence:            DABACABA  
Display Prelim Rslt?            Yes  
Display Third Digit?            Yes  
Inhib Printer(Y/N)?            No  
Display Volume?                No  
Disable On Memfull?            Yes  
# of Print Copies?              1  
Select Std (D/W/I)?            Dry  
Standard Value?                0.080  
Standard Lot #?                 02621080A1  
Standard Cyl #?                 2  
Standard Expiration?            03/05/2023  
Oper No?                        132128

Flow Cal. Date:                 03/30/2022  
Slope                            662  
Intercept                        -676090

IR Calibration Date:            03/30/2022  
                                  3um                            9um

-----  
0th Coef (\*100):                -20151                        -12325  
1st Coef (\*100):                267293                        130383  
2nd Coef (\*100):                4119                            1500  
H2O adj (mg/l\*10k):            550                             327

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003069  
Location = TOXL      8164.14.00 09/16  
06/28/2022      15:10

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:10
02 Std. Sol.	0.018	15:11
03 Room Air	0.000	15:12
04 Std. Sol.	0.018	15:12
05 Room Air	0.000	15:13
06 Std. Sol.	0.018	15:14
07 Room Air	0.000	15:14

08 Sim Temp = 34.0°C

Simul Ser No = MP3070  
Std Sol No = 22CP41184  
County = 08      Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Low Ac Check 0.020*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003069  
Location = TOXL                    8164.14.00 09/16  
06/28/2022                            16:04

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:05
02 Std. Sol.	0.395	16:06
03 Room Air	0.000	16:07
04 Std. Sol.	0.398	16:08
05 Room Air	0.000	16:08
06 Std. Sol.	0.398	16:09
07 Room Air	0.000	16:09

08 Sim Temp = 34.0°C

Simul Ser No = MP3065  
Std Sol No = 202103E  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *High AC check 0.400*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003069  
Location = TOXL                    8164.14.00 09/16  
06/28/2022                            16:17

Test	AC	Time
01 Room Air	0.000	16:18
02 *Subject Test	INT*	16:19
03 Room Air	0.000	16:20

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT CK  
Dr. Lic. = ND/DIS821456  
Lot No = 34921080A1  
Cyl No = 44  
Expiration Date = 02/05/2024  
County = 08                            Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks:

*RFI at 9:00*  
*Interferent check*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003069  
Location = TOXL                    8164.14.00 09/16  
06/28/2022                            16:20

Test	AC	Time
01 Diagnostic	OK	16:21
02 Room Air	0.000	16:22
03 *Subject Test	RFI*	16:22
04 Room Air	0.000	16:23

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHK  
Dr. Lic. = ND/DIS821456  
Lot No = 34921080A1  
Cyl No = 44  
Expiration Date = 02/05/2024  
County = 08                            Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *RFI Check*




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003069  
Location = TOXL                    8164.14.00 09/16  
06/28/2022                            16:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:25
02 Std. Gas	0.079	16:25
03 Room Air	0.000	16:25
04 Std. Gas	0.079	16:26
05 Room Air	0.000	16:26
06 Std. Gas	0.079	16:27
07 Room Air	0.000	16:27

Lot No = 34921080A1  
Cyl No = 44  
Exp Date = 02/05/2024  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 1-3*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003069  
Location = TOXL                    8164.14.00 09/16  
06/28/2022                            16:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:31
02 Std. Gas	0.079	16:31
03 Room Air	0.000	16:32
04 Std. Gas	0.079	16:32
05 Room Air	0.000	16:32
06 Std. Gas	0.079	16:33
07 Room Air	0.000	16:33

Lot No = 34921080A1  
Cyl No = 44  
Exp Date = 02/05/2024  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 4-6*




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003069  
Location = TOXL                    8164.14.00 09/16  
06/28/2022                            16:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:36
02 Std. Gas	0.079	16:36
03 Room Air	0.000	16:37
04 Std. Gas	0.079	16:37
05 Room Air	0.000	16:38
06 Std. Gas	0.079	16:38
07 Room Air	0.000	16:39

Lot No = 34921080A1  
Cyl No = 44  
Exp Date = 02/05/2024  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry calibration check # 7-9*