

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-003058

Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 419 psi Regulator: 450 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP5290 Lot #: 21340 Exp. Date: 8/24/23
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP3067 Lot #: 21420 Exp. Date: 10/20/23

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR 7345 Lot #: 1657 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.08 AC  
Gas Cyl Lot #: 02621080A1 Cyl #: 31 Exp. Date: 3/5/23

Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.080</u>	Test 5 <u>0.078</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.080</u>
Average <u>0.079</u>		

C. Remarks/Maintenance: Annual inspection

Replaced breath tube + simulator return o-rings.

Instrument is acceptable to be used in the field.

Janelle Pertschiller  
Breath Analyst Signature

6/29/2022  
Date

[Signature]  
Reviewed by

063022  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003058  
Location = TOXL                 8164.14.00 09/16  
06/29/2022                        09:45

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:                Enabled  
Start Test Sequence:            DABACABA  
Display Prelim Rslt?            Yes  
Display Third Digit?            Yes  
Inhib Printer(Y/N)?            No  
Display Volume?                No  
Disable On Memfull?            Yes  
# of Print Copies?              1  
Select Std (D/W/I)?            Dry  
Standard Value?                0.080  
Standard Lot #?                 24119080A1  
Standard Cyl #?                 5  
Standard Expiration?            11/05/2021  
Oper No?                        132128

Flow Cal. Date:                 04/07/2008  
Slope                            698  
Intercept                       -678970

IR Calibration Date:            04/15/2015  
                                    3um                            9um

-----  
0th Coef(\*100):                -18934                        -10185  
1st Coef(\*100):                272557                        131040  
2nd Coef(\*100):                2331                            1317  
H2O adj(mg/l\*10k):            333                             280

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003058  
Location = TOXL                    8164.14.00 09/16  
06/29/2022                            09:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:50
02 Std. Sol.	0.018	09:51
03 Room Air	0.000	09:51
04 Std. Sol.	0.019	09:52
05 Room Air	0.000	09:53
06 Std. Sol.	0.019	09:53
07 Room Air	0.000	09:54

08 Sim Temp = 34.0°C

Simul Ser No = MP5290  
Std Sol No = 21340  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003058  
Location = TOXL                    8164.14.00 09/16  
06/29/2022                            09:56

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:56
02 Std. Sol.	0.292	09:57
03 Room Air	0.000	09:57
04 Std. Sol.	0.294	09:58
05 Room Air	0.000	09:59
06 Std. Sol.	0.296	09:59
07 Room Air	0.000	10:00

08 Sim Temp = 34.0°C

Simul Ser No = MP3067  
Std Sol No = 21420  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *High AC check 0.300*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003058  
Location = TOXL                8164.14.00 09/16  
06/29/2022                        10:04

Test	AC	Time
01 Room Air	0.000	10:05
02 *Subject Test	INT*	10:07
03 Room Air	0.000	10:08

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                    Weight = NA  
Test = DUI                        Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 02621080A1  
Cyl No = 31  
Expiration Date = 03/05/2023  
County = 08                        Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003058  
Location = TOXL                    8164.14.00 09/16  
06/29/2022                            10:11

Test	AC	Time
01 Diagnostic	OK	10:11
02 Room Air	0.000	10:12
03 *Subject Test	RFI*	10:13
04 Room Air	0.000	10:13

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                      Weight = NA  
Test = DUI                            Cit = RFI CHK  
Dr. Lic. = ND/TES989643  
Lot No = 02621080A1  
Cyl No = 31  
Expiration Date = 03/05/2023  
County = 08                          Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *RFI Check*




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003058  
Location = TOXL                    8164.14.00 09/16  
06/29/2022                            10:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:30
02 Std. Gas	0.079	10:30
03 Room Air	0.000	10:30
04 Std. Gas	0.080	10:31
05 Room Air	0.000	10:31
06 Std. Gas	0.079	10:32
07 Room Air	0.000	10:32

Lot No = 02621080A1  
Cyl No = 31  
Exp Date = 03/05/2023  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check # 1-3*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003058  
Location = TOXL                    8164.14.00 09/16  
06/29/2022                            10:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:36
02 Std. Gas	0.079	10:37
03 Room Air	0.000	10:37
04 Std. Gas	0.078	10:38
05 Room Air	0.000	10:38
06 Std. Gas	0.079	10:38
07 Room Air	0.000	10:39

Lot No = 02621080A1  
Cyl No = 31  
Exp Date = 03/05/2023  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check 4-6*




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003058  
Location = TOXL      8164.14.00 09/16  
06/29/2022      10:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:41
02 Std. Gas	0.079	10:41
03 Room Air	0.000	10:42
04 Std. Gas	0.079	10:42
05 Room Air	0.000	10:43
06 Std. Gas	0.080	10:43
07 Room Air	0.000	10:44

Lot No = 02621080A1  
Cyl No = 31  
Exp Date = 03/05/2023  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #7-9*