

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00491

Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 419 psi Regulator: 400 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
    - a.  Low AC (Level 1,C): Known Value  $\leq$  0.03 AC: 0.020 AC  
Sim. Ser #: MP5290 Lot #: 21340 Exp. Date: 8/24/23
    - b.  High AC (Level 1,C): Known Value  $\geq$  0.25 AC: 0.300 AC  
Sim. Ser #: MP3067 Lot #: 21420 Exp. Date: 10/20/23

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR7345 Lot #: 1657 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.08 AC  
Gas Cyl Lot #: 02621080A1 Cyl #: 31 Exp. Date: 3/5/23

Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.080</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.079</u>
Average <u>0.080</u>		

C. Remarks/Maintenance: Annual inspection  
Replaced breath tube + simulator return o-rings.

Instrument is acceptable to be used in the field.

Janelle Portschiller  
Breath Analyst Signature

06/29/2022  
Date

[Signature]  
Reviewed by

063022  
Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004191  
Location = TOXL      8164.14.00 09/16  
06/29/2022      13:03

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:03
02 Std. Sol.	0.018	13:04
03 Room Air	0.000	13:05
04 Std. Sol.	0.019	13:06
05 Room Air	0.000	13:06
06 Std. Sol.	0.020	13:07
07 Room Air	0.000	13:08

08 Sim Temp = 34.0°C

Simul Ser No = MP5290

Std Sol No = 21340

County = 08

Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004191  
Location = TOXL      8164.14.00 09/16  
06/29/2022      13:10

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:10
02 Std. Sol.	0.297	13:11
03 Room Air	0.000	13:12
04 Std. Sol.	0.297	13:13
05 Room Air	0.000	13:13
06 Std. Sol.	0.297	13:14
07 Room Air	0.000	13:15

08 Sim Temp = 34.0°C

Simul Ser No = MP3067

Std Sol No = 21420

County = 08

Oper No. = 132128

*Janelle Portscheller*

Operator Signature  
JANELLE PORTSCHELLER

Remarks: *High Ac check 0.300*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004191  
Location = TOXL                    8164.14.00 09/16  
06/29/2022    13:16

Test	AC	Time
01 Room Air	0.000	13:17
02 *Subject Test	INT*	13:18
03 Room Air	0.000	13:19

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                                    Weight = NA  
Test = DUI                                    Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 02621080A1  
Cyl No = 31  
Expiration Date = 03/05/2023  
County = 08                                    Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004191  
Location = TOXL                    8164.14.00 09/16  
06/29/2022                            13:22

Test	AC	Time
01 Diagnostic	OK	13:23
02 Room Air	0.000	13:23
03 *Subject Test	RFI*	13:23
04 Room Air	0.000	13:24

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                            Weight = NA  
Test = DUI                                    Cit = RFI CHK  
Dr. Lic. = ND/TES989643  
Lot No = 02621080A1  
Cyl No = 31  
Expiration Date = 03/05/2023  
County = 08                                    Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *RFI check*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004191  
Location = TOXL      8164.14.00 09/16  
06/29/2022      13:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:25
02 Std. Gas	0.080	13:25
03 Room Air	0.000	13:26
04 Std. Gas	0.080	13:26
05 Room Air	0.000	13:27
06 Std. Gas	0.080	13:27
07 Room Air	0.000	13:28

Lot No = 02621080A1  
Cyl No = 31  
Exp Date = 03/05/2023  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004191  
Location = TOXL                    8164.14.00 09/16  
06/29/2022                            13:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:33
02 Std. Gas	0.080	13:34
03 Room Air	0.000	13:34
04 Std. Gas	0.080	13:35
05 Room Air	0.000	13:35
06 Std. Gas	0.080	13:36
07 Room Air	0.000	13:36

Lot No = 02621080A1  
Cyl No = 31  
Exp Date = 03/05/2023  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3 & 4-6*




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004191  
Location = TOXL      8164.14.00 09/16  
06/29/2022      13:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:37
02 Std. Gas	0.080	13:38
03 Room Air	0.000	13:38
04 Std. Gas	0.080	13:38
05 Room Air	0.000	13:39
06 Std. Gas	0.079	13:39
07 Room Air	0.000	13:40

Lot No = 02621080A1  
Cyl No = 31  
Exp Date = 03/05/2023  
County = 08      Oper No. = 132128

  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #7-9*