

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005940

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 915 psi Regulator: 900 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: MP3070 Lot #: 22CP41184 Exp. Date: 3/1/24
 - b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.400 AC
Sim. Ser #: MP 3065 Lot #: 202103E Exp. Date: 3/24/23

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: 1CS6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 34921080A1 Cyl #: 44 Exp. Date: 2/5/24

Test 1 <u>0.079</u>	Test 4 <u>0.078</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.078</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.078</u>
Average <u>0.079</u>		

C. Remarks/Maintenance: Annual inspection.

Replaced breath tube + simulator return O-rings.
Added electrical tape to breath tube hose to prevent wear.
Replaced external battery.

Instrument is acceptable to be used in the field.

Jennifer Perichella
Breath Analyst Signature

06/29/2022
Date

Reviewed by

OC3022
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
06/29/2022 13:28

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:29
02 Std. Sol.	0.393	13:30
03 Room Air	0.000	13:30
04 Std. Sol.	0.396	13:31
05 Room Air	0.000	13:31
06 Std. Sol.	0.397	13:32
07 Room Air	0.000	13:33


08 Sim Temp = 34.0°C

Simul Ser No = MP3065

Std Sol No = 202103E

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High AC check 0.400*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
06/29/2022 13:47

Test	AC	Time
01 Diagnostic	OK	13:48
02 Room Air	0.000	13:48
03 *Subject Test	RFI*	13:49
04 Room Air	0.000	13:49

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS821456
Lot No = 34921080A1
Cyl No = 44
Expiration Date = 02/05/2024
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI Check*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
06/29/2022 13:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:51
02 Std. Gas	0.079	13:51
03 Room Air	0.000	13:52
04 Std. Gas	0.079	13:52
05 Room Air	0.000	13:52
06 Std. Gas	0.079	13:53
07 Room Air	0.000	13:53

Lot No = 34921080A1
Cyl No = 44
Exp Date = 02/05/2024
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
06/29/2022 13:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:55
02 Std. Gas	0.078	13:55
03 Room Air	0.000	13:56
04 Std. Gas	0.078	13:56
05 Room Air	0.000	13:57
06 Std. Gas	0.079	13:57
07 Room Air	0.000	13:57

Lot No = 34921080A1
Cyl No = 44
Exp Date = 02/05/2024
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-6*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
06/29/2022 13:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:59
02 Std. Gas	0.079	13:59
03 Room Air	0.000	14:00
04 Std. Gas	0.079	14:00
05 Room Air	0.000	14:01
06 Std. Gas	0.078	14:01
07 Room Air	0.000	14:02


Lot No = 34921080A1

Cyl No = 44

Exp Date = 02/05/2024

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry calibration check # 7-9*