

**INTOXILYZER® 8000 INSPECTION**

Intoxilyzer® 8000 Serial Number: 80-004204 Location: TOXL

A. General Setup and Checks:

- 1.  Diagnostic Tests Pass and Instrument in Ready Mode
- 2.  Breath tube heated
- 3.  Date, time and location code (Level 2,E). Re-set if necessary.
- 4.  Print test (Level 1,P). Sign and attach test record.
- 5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 733 psi Regulator: 750 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
  - 1. Wet Calibration Check
    - a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP 6036 Lot #: 21340 Exp. Date: 8/24/23
    - b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP 3068 Lot #: 202201F Exp. Date: 1/18/24

Configure dry gas standard for the following tests (Level 1,S).

- 2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR 3847 Lot #: 1CS6 Exp. Date: NA  
 Display reads "Interferent Detect"
- 3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
- 4.  Dry Calibration Check (Level 1,C): Known Value 0.08 AC  
Gas Cyl Lot #: 34921080A1 Cyl #: 47 Exp. Date: 2/5/24

Test 1 <u>0.080</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.079</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.078</u>	Test 9 <u>0.079</u>
Average <u>0.079</u>		

C. Remarks/Maintenance: Annual inspection.  
Replaced breath tube and simulator return O-rings.

Instrument is acceptable to be used in the field.

Janella Putschiller  
Breath Analyst Signature

06/30/2022  
Date

[Signature]  
Reviewed by

070122  
Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004204  
Location = TOXL      8164.14.00 09/16  
06/30/2022      10:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:56
02 Std. Sol.	0.017	10:56
03 Room Air	0.000	10:57
04 Std. Sol.	0.017	10:58
05 Room Air	0.000	10:58
06 Std. Sol.	0.017	10:59
07 Room Air	0.000	10:59

08 Sim Temp = 34.0°C

Simul Ser No = MP6036  
Std Sol No = 21340  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Low Ac check 0.020*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                    8164.14.00 09/16  
06/30/2022                            11:01

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:01
02 Std. Sol.	0.291	11:02
03 Room Air	0.000	11:03
04 Std. Sol.	0.293	11:03
05 Room Air	0.000	11:04
06 Std. Sol.	0.293	11:05
07 Room Air	0.000	11:05

08 Sim Temp = 34.0°C

Simul Ser No = MP3068  
Std Sol No = 202201F  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *High Ac check 0.300*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                    8164.14.00 09/16  
06/30/2022                            11:11

Test	AC	Time
01 Room Air	0.000	11:11
02 *Subject Test	INT*	11:12
03 Room Air	0.000	11:13

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT CK  
Dr. Lic. = ND/DIS921456  
Lot No = 34921080A1  
Cyl No = 47  
Expiration Date = 02/05/2024  
County = 08                            Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks:

~~RFI~~ check <sup>JP</sup>  
Interferent check

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                 8164.14.00 09/16  
06/30/2022                        11:20

Test	AC	Time
01 Diagnostic	OK	11:21
02 Room Air	0.000	11:21
03 *Subject Test	RFI*	11:21
04 Room Air	0.000	11:22

\*Invalid Test  
  Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHK  
Dr. Lic. = ND/DIS921456  
Lot No = 34921080A1  
Cyl No = 47  
Expiration Date = 02/05/2024  
County = 08                            Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *RFI check*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004204  
Location = TOXL      8164.14.00 09/16  
06/30/2022      11:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:24
02 Std. Gas	0.080	11:24
03 Room Air	0.000	11:25
04 Std. Gas	0.079	11:25
05 Room Air	0.000	11:26
06 Std. Gas	0.079	11:26
07 Room Air	0.000	11:26

Lot No = 34921080A1  
Cyl No = 47  
Exp Date = 02/05/2024  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                    8164.14.00 09/16  
06/30/2022                            11:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:27
02 Std. Gas	0.079	11:28
03 Room Air	0.000	11:28
04 Std. Gas	0.079	11:28
05 Room Air	0.000	11:29
06 Std. Gas	0.078	11:29
07 Room Air	0.000	11:30


Lot No = 34921080A1

Cyl No = 47

Exp Date = 02/05/2024

County = 08

Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-6*




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004204  
Location = TOXL      8164.14.00 09/16  
06/30/2022      11:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:32
02 Std. Gas	0.079	11:32
03 Room Air	0.000	11:33
04 Std. Gas	0.079	11:33
05 Room Air	0.000	11:33
06 Std. Gas	0.079	11:34
07 Room Air	0.000	11:34

Lot No = 34921080A1  
Cyl No = 47  
Exp Date = 02/05/2024  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #7-9*