

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-604201 Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 733 psi Regulator: 750 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: MP3060 Lot #: 202201A Exp. Date: 1/11/24
- b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.300 AC
Sim. Ser #: MP3064 Lot #: 202012A Exp. Date: 12/8/22

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7345 Lot #: 1CS7 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.08 AC
Gas Cyl Lot #: 34921080A1 Cyl #: 49 Exp. Date: 2/5/24

Test 1	<u>0.080</u>	Test 4	<u>0.079</u>	Test 7	<u>0.079</u>
Test 2	<u>0.079</u>	Test 5	<u>0.079</u>	Test 8	<u>0.079</u>
Test 3	<u>0.079</u>	Test 6	<u>0.079</u>	Test 9	<u>0.079</u>
Average	<u>0.079</u>				

C. Remarks/Maintenance: Annual inspection.

Replace breath tube + simulator return o-rings.

Instrument is acceptable to be used in the field.

Janelle Pertschiller
Breath Analyst Signature

06/30/2022
Date

[Signature]
Reviewed by

07/20/22
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004201
Location = TOXL 8164.14.00 09/16
06/30/2022 10:46

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 24119080A1
Standard Cyl #? 5
Standard Expiration? 11/05/2021
Oper No? 132128

Flow Cal. Date: 04/25/2018
Slope 580
Intercept -164747

IR Calibration Date: 04/26/2018
 3um 9um

0th Coef(*100): -26833 -28328
1st Coef(*100): 272526 132392
2nd Coef(*100): 3966 1397
H2O adj(mg/l*10k): 561 514

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004201
Location = TOXL 8164.14.00 09/16
06/30/2022 10:52

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:52
02 Std. Sol.	0.019	10:53
03 Room Air	0.000	10:54
04 Std. Sol.	0.019	10:55
05 Room Air	0.000	10:55
06 Std. Sol.	0.019	10:56
07 Room Air	0.000	10:56

08 Sim Temp = 34.0°C

Simul Ser No = MP3060
Std Sol No = 202201A
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004201
Location = TOXL 8164.14.00 09/16
06/30/2022 10:58

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:58
02 Std. Sol.	0.299	10:59
03 Room Air	0.000	11:00
04 Std. Sol.	0.302	11:01
05 Room Air	0.000	11:01
06 Std. Sol.	0.303	11:02
07 Room Air	0.000	11:02


08 Sim Temp = 34.0°C

Simul Ser No = MP3064

Std Sol No = 202012A

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High AC check 0.300*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004201
Location = TOXL 8164.14.00 09/16
06/30/2022 11:05

Test	AC	Time
01 Room Air	0.000	11:06
02 *Subject Test	INT*	11:06
03 Room Air	0.000	11:07

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS821456
Lot No = 34921080A1
Cyl No = 49
Expiration Date = 02/05/2024
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent Check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004201
Location = TOXL 8164.14.00 09/16
06/30/2022 11:15

Test	AC	Time
01 Diagnostic	OK	11:16
02 Room Air	0.000	11:17
03 *Subject Test	RFI*	11:17
04 Room Air	0.000	11:17

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS821456
Lot No = 34921080A1
Cyl No = 49
Expiration Date = 02/05/2024
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI Check*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004201
Location = TOXL 8164.14.00 09/16
06/30/2022 11:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:22
02 Std. Gas	0.080	11:22
03 Room Air	0.000	11:23
04 Std. Gas	0.079	11:23
05 Room Air	0.000	11:24
06 Std. Gas	0.079	11:24
07 Room Air	0.000	11:25

Lot No = 34921080A1
Cyl No = 49
Exp Date = 02/05/2024
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004201
Location = TOXL 8164.14.00 09/16
06/30/2022 11:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:25
02 Std. Gas	0.079	11:26
03 Room Air	0.000	11:26
04 Std. Gas	0.079	11:27
05 Room Air	0.000	11:27
06 Std. Gas	0.079	11:28
07 Room Air	0.000	11:28

Lot No = 34921080A1
Cyl No = 49
Exp Date = 02/05/2024
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 4-6*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004201
Location = TOXL 8164.14.00 09/16
06/30/2022 11:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:31
02 Std. Gas	0.079	11:31
03 Room Air	0.000	11:32
04 Std. Gas	0.079	11:32
05 Room Air	0.000	11:33
06 Std. Gas	0.079	11:33
07 Room Air	0.000	11:34

Lot No = 34921080A1
Cyl No = 49
Exp Date = 02/05/2024
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration check #1-3 per 7-9*