

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004196

Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 380 psi Regulator: 400 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP5290 Lot #: 21340 Exp. Date: 8/24/23
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP3067 Lot #: 21420 Exp. Date: 10/20/23

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DL7345 Lot #: 1CS7 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.08 AC  
Gas Cyl Lot #: 0262108041 Cyl #: 31 Exp. Date: 3/5/23

Test 1 <u>0.078</u>	Test 4 <u>0.078</u>	Test 7 <u>0.077</u>
Test 2 <u>0.077</u>	Test 5 <u>0.077</u>	Test 8 <u>0.078</u>
Test 3 <u>0.078</u>	Test 6 <u>0.077</u>	Test 9 <u>0.077</u>
Average <u>0.077</u>		

C. Remarks/Maintenance: Annual inspection.

Replaced breath tube + simulator return O-rings.

Instrument is acceptable to be used in the field.

Janelle Perichillo  
Breath Analyst Signature

06/30/2022  
Date

[Signature]  
Reviewed by

070122  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004196  
Location = TOXL                 8164.14.00 09/16  
06/30/2022                        10:46

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:                Enabled  
Start Test Sequence:            DABACABA  
Display Prelim Rslt?            Yes  
Display Third Digit?            Yes  
Inhib Printer(Y/N)?            No  
Display Volume?                No  
Disable On Memfull?            Yes  
# of Print Copies?              1  
Select Std (D/W/I)?            Dry  
Standard Value?                0.080  
Standard Lot #?                 02621080A1  
Standard Cyl #?                26  
Standard Expiration?            03/05/2023  
Oper No?                        132128

Flow Cal. Date:                 04/15/2020  
Slope                            657  
Intercept                        -683120

IR Calibration Date:            04/16/2020  
                                    3um                            9um

-----  
0th Coef(\*100):                -35275                        -26830  
1st Coef(\*100):                271852                        134658  
2nd Coef(\*100):                1067                            849  
H2O adj(mg/l\*10k):            413                             302

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *print test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004196  
Location = TOXL      8164.14.00 09/16  
06/30/2022      10:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:53
02 Std. Sol.	0.018	10:54
03 Room Air	0.000	10:55
04 Std. Sol.	0.018	10:55
05 Room Air	0.000	10:56
06 Std. Sol.	0.018	10:57
07 Room Air	0.000	10:57

08 Sim Temp = 34.0°C

Simul Ser No = MP5290  
Std Sol No = 21340  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Low Ac check 0.020*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004196  
Location = TOXL      8164.14.00 09/16  
06/30/2022      10:59

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:59
02 Std. Sol.	0.295	11:00
03 Room Air	0.000	11:01
04 Std. Sol.	0.297	11:01
05 Room Air	0.000	11:02
06 Std. Sol.	0.298	11:03
07 Room Air	0.000	11:03

08 Sim Temp = 34.0°C

Simul Ser No = MP3067  
Std Sol No = 21420  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *High Ac Check 0.300*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004196  
Location = TOXL                    8164.14.00 09/16  
06/30/2022                            11:07

Test	AC	Time
01 Room Air	0.000	11:08
02 *Subject Test	INT*	11:08
03 Room Air	0.000	11:09

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                      Weight = NA  
Test = DUI                          Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 02621080A1  
Cyl No = 31  
Expiration Date = 03/05/2023  
County = 08                          Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Interferent Check*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004196  
Location = TOXL                    8164.14.00 09/16  
06/30/2022                            11:16

Test	AC	Time
01 Diagnostic	OK	11:17
02 Room Air	0.000	11:17
03 *Subject Test	RFI*	11:19
04 Room Air	0.000	11:20

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                      Weight = NA  
Test = DUI                            Cit = RFI CHK  
Dr. Lic. = ND/TES989643  
Lot No = 02621080A1  
Cyl No = 31  
Expiration Date = 03/05/2023  
County = 08                          Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *RFI check*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004196  
Location = TOXL                    8164.14.00 09/16  
06/30/2022                            11:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:22
02 Std. Gas	0.078	11:23
03 Room Air	0.000	11:23
04 Std. Gas	0.077	11:24
05 Room Air	0.000	11:24
06 Std. Gas	0.078	11:24
07 Room Air	0.000	11:25

Lot No = 02621080A1  
Cyl No = 31  
Exp Date = 03/05/2023  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry calibration check #1-3*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004196  
Location = TOXL                    8164.14.00 09/16  
06/30/2022                            11:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:26
02 Std. Gas	0.078	11:26
03 Room Air	0.000	11:27
04 Std. Gas	0.077	11:27
05 Room Air	0.000	11:27
06 Std. Gas	0.077	11:28
07 Room Air	0.000	11:28

Lot No = 02621080A1  
Cyl No = 31  
Exp Date = 03/05/2023  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-6*




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004196  
Location = TOXL      8164.14.00 09/16  
06/30/2022      11:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:31
02 Std. Gas	0.077	11:31
03 Room Air	0.000	11:32
04 Std. Gas	0.078	11:32
05 Room Air	0.000	11:33
06 Std. Gas	0.077	11:33
07 Room Air	0.000	11:34

Lot No = 02621080A1  
Cyl No = 31  
Exp Date = 03/05/2023  
County = 08      Oper No. = 132128

  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 7-9*