



**ETHANOL GAS STANDARD CYLINDER REPORT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Nikko, Jacob, V</i> <i>132597</i>		
Location <i>Grafton PD</i>	Intoxilyzer® Serial Number <i>80-003067</i>	
Gas Lot Number <i>02621080A1</i>	Gas Cylinder Number <i>18</i>	Gas Expiration Date <i>3/5/2023</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. *0.028* AC
  - B. *0.028* AC
  - C. *0.029* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Jacob Nikko</i>	Date <i>12/20/2022</i>
Reviewed By (Crime Laboratory Use Only) <i>Janelle Pertschiller</i>	Date <i>22 Dec 2022</i>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003067  
Location = GFTN      8164.14.00 09/16  
12/20/2022      15:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:33
02 Std. Gas	0.078	15:34
03 Room Air	0.000	15:34
04 Std. Gas	0.078	15:34
05 Room Air	0.000	15:35
06 Std. Gas	0.079	15:35
07 Room Air	0.000	15:36

Lot No = 02621080A1  
Cyl No = 18  
Exp Date = 03/05/2023  
County = 50      Oper No. = 132597

  
\_\_\_\_\_  
Operator Signature  
JACOB NIKKO

Remarks: