



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

| | | |
|---|--|--|
| Chemical Test Operator Name (Print) <i>Michael Hayden (NOHP-237)</i> | | |
| Location <i>Napoleon (NAP0)</i> | Intoxilyzer® Serial Number <i>80-004200</i> | |
| Gas Lot Number <i>34921080A2</i> | Gas Cylinder Number <i>059</i> | Gas Expiration Date <i>02/05/2024</i> |

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.081* AC
 B. *0.082* AC
 C. *0.082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

| | |
|---|---------------------------|
| Chemical Test Operator Signature <i>[Signature]</i> | Date <i>01/06/2023</i> |
| Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i> | Date <i>09Jan2023</i> |

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = NAPO 8164.14.00 09/16
01/06/2023 17:18

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 17:19 |
| 02 Std. Gas | 0.081 | 17:19 |
| 03 Room Air | 0.000 | 17:20 |
| 04 Std. Gas | 0.082 | 17:20 |
| 05 Room Air | 0.000 | 17:21 |
| 06 Std. Gas | 0.082 | 17:21 |
| 07 Room Air | 0.000 | 17:22 |

Lot No = 34921080A1
Cyl No = 59
Exp Date = 02/05/2024
County = 24 Oper No. = 130163



Operator Signature
MICHAEL HAYEN

Remarks:

Form 106-I8000