



ETHANOL GAS STANDARD CYLINDER REPORT
NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (08/2018)

Chemical Test Operator Name (Print) DANA R MIDDLETON		
Location NEW SAGEN (MORTON COUNTY)	Intoxilyzer® Serial Number 80-004191	
Gas Lot Number 01923080A3	Gas Cylinder Number 24	Gas Expiration Date 02-05-2025

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S)
- 2. Perform an ACA Test (Level 1, Function C)
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. **083** AC
 - B. **083** AC
 - C. **083** AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G)
 - B. ACA Test Record

Chemical Test Operator Signature <i>Dana R Middleton</i>	Date 2/27/23
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date 27Feb2023


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = NS 8164.14.00 09/16
02/27/2023 00:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	00:07
02 Std. Gas	0.083	00:07
03 Room Air	0.000	00:08
04 Std. Gas	0.083	00:08
05 Room Air	0.000	00:09
06 Std. Gas	0.083	00:10
07 Room Air	0.000	00:10

Lot No = 01923080A3
Cyl No = 24
Exp Date = 02/05/2025
County = 30 Oper No. = 079706



Operator Signature
DANA MIDDLETON

Remarks:

Form 106-I8000