



# ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL



CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <b>Kari Stuart</b>		
Location <b>1201 12th St. SE, Watford City, ND 58095A</b>	Intoxilyzer® Serial Number <b>000-005302</b>	
Gas Lot Number <b>33321000AA</b>	Gas Cylinder Number <b>012</b>	Gas Expiration Date <b>11/5/24</b>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. 0.001 AC
  - B. 0.002 AC
  - C. 0.001 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature 	Date <b>4/7/22</b>
Reviewed By (Crime Laboratory Use Only) 	Date <b>08/12/22</b>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005362  
Location = MCSO      8164.14.00 09/16  
04/07/2022      07:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	07:47
02 Std. Gas	0.081	07:47
03 Room Air	0.000	07:48
04 Std. Gas	0.082	07:48
05 Room Air	0.000	07:49
06 Std. Gas	0.081	07:49
07 Room Air	0.000	07:49

Lot No = 33321080A4  
Cyl No = 12  
Exp Date = 01/05/2024  
County = 27      Oper No. = 132007

  
Operator Signature  
KARI STUART


Remarks:  
ACA test

Form 106-I8000

STATE OF NORTH DAKOTA }  
COUNTY OF MCKENZIE }<sup>ss</sup>

I certify that the foregoing is a true and correct  
copy of the original document on file in this office.  
Dated at Watford City, North Dakota.

on this 7th day of April, 2022

  
McKenzie County Sheriff's Department