



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) Shannon Liebold		
Location Oliver County Sheriff's Department		Intoxilyzer® Serial Number 80-005958
Gas Lot Number 34921080A1	Gas Cylinder Number 58	Gas Expiration Date 02/05/2024

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0.079 AC
 - B. 0.079 AC
 - C. 0.080 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Shannon Liebold</i>	Date 04/12/2022
Reviewed By (Crime Laboratory Use Only) <i>Janelle Poutschiller</i>	Date 20 Apr 22


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005958
Location = OLIV 8164.14.00 09/16
04/12/2022 17:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:52
02 Std. Gas	0.079	17:52
03 Room Air	0.000	17:53
04 Std. Gas	0.079	17:53
05 Room Air	0.000	17:54
06 Std. Gas	0.080	17:54
07 Room Air	0.000	17:55

Lot No = 34921080A1
Cyl No = 58
Exp Date = 02/05/2024
County = 33 Oper No. = 020904


Operator Signature
SHANNON LIEBOLD

Remarks: