




INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-004947	Instrument Location Dickinson LEC
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	


Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature  #4649	Date 12/8/22
--	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By 	Certified Date 13 Dec 2022
--	--------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = DKPD 8164.14.00 09/16
12/08/2022 15:24

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34921080A1
Standard Cyl #? 040
Standard Expiration? 02/05/2024
Oper No? 130935

Flow Cal. Date: 04/07/2020
Slope 618
Intercept -565524

IR Calibration Date: 04/07/2020
 3um 9um

0th Coef(*100):	-22371	-26151
1st Coef(*100):	260822	133081
2nd Coef(*100):	1936	1144
H2O adj(mg/l*10k):	495	535

***** Printer Test End *****


Operator Signature
KEVIN ELDRIDGE

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = DKPD 8164.14.00 09/16
12/08/2022 15:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:26
02 Std. Gas	0.080	15:26
03 Room Air	0.000	15:26
04 Std. Gas	0.080	15:27
05 Room Air	0.000	15:27
06 Std. Gas	0.081	15:28
07 Room Air	0.000	15:28

Lot No = 34921080A1
Cyl No = 040
Exp Date = 02/05/2024
County = 45 Oper No. = 130935

 #4649
Operator Signature
KEVIN ELDRIDGE

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = DKPD 8164.14.00 09/16
12/08/2022 15:29

Test	AC	Time
01 Room Air	RFI*	15:30
02 Room Air	RFI*	15:30

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST NMN
Sub DOB = 01/01/1975
Sub Sex = Male Weight = 200
Test = OTH Cit = N/A
Dr. Lic. = ND/TES755555
Lot No = 34921080A1
Cyl No = 040
Expiration Date = 02/05/2024
County = 45 Oper No. = 130935


Operator Signature
KEVIN ELDRIDGE

Remarks: