



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	William R Petty				
Location:	GF AF				
Intoxilyzer® Serial Number:	80-004949				
Gas Lot Number:	659358	Gas Cylinder Number:	39	Expiration Date:	09/03/2011
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed				
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: 0.076 %					
Test 2: 0.077 %					
Test 3: 0.077 %					
(ATTACH TEST RECORD)					

Will Petty  
Chemical Test Operator Signature

16 Aug 2010  
Date

Deb Kashur  
Reviewed By

22 Mar 2011  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004949  
Location = GFAP      8164.13.00 06/09  
08/16/2010      08:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:25
02 Std. Gas	0.076	08:25
03 Room Air	0.000	08:26
04 Std. Gas	0.077	08:26
05 Room Air	0.000	08:27
06 Std. Gas	0.077	08:27
07 Room Air	0.000	08:28

Lot No = 659358  
Cyl No = 39  
Exp Date = 09/03/2011  
County = 18      Oper No. = 120011



Operator Signature  
WILLIAM R. PETTY

Remarks:

Form 106-I8000

80-004949  
P. 97 of 97  
km