



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Lt. Tracy L. Meidinger #258</u>		
Location: <u>UND-PO</u>		
Intoxilyzer® Serial Number: <u>80-004934</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>31</u>	Expiration Date: <u>09-30-2011</u> <sup>03</sup> <sub>pk</sub>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>082</u> % Test 2: 0. <u>082</u> % Test 3: 0. <u>081</u> %		
(ATTACH TEST RECORD)		

Lt. Tracy L. Meidinger  
 Chemical Test Operator Signature

03-04-2011  
 Date

Deb Kashner  
 Reviewed By

23 Mar 2011  
 Date

80-004934  
 P. 67 of 68  
 RM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004934  
Location = UNDP                    8164.13.00 06/09  
03/04/2011                            14:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:43
02 Std. Gas	0.082	14:44
03 Room Air	0.000	14:44
04 Std. Gas	0.082	14:45
05 Room Air	0.000	14:45
06 Std. Gas	0.081	14:45
07 Room Air	0.000	14:46

Lot No = 659358  
Cyl No = 31  
Exp Date = 09/03/2011  
County = 18                            Oper No. = 119409

*Tracy L. Meidinger #258*  
Operator Signature  
TRACY L. MEIDINGER

Remarks:

Form 106-I8000

80-004934  
P. 68 of 68  
FM