



ETHANOL GAS STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Darcy W Aberle</i>		
Location <i>Williston LEC</i>	Intoxilyzer® Serial Number <i>80-005946</i>	
Gas Lot Number <i>09216080A3</i>	Gas Cylinder Number <i>44</i>	Gas Expiration Date <i>05/05/2018</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.081* AC
 - B. *0.082* AC
 - C. *0.083* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Darcy W Aberle</i>	Date <i>09-28-16</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Shanaver</i>	Date <i>28 Sept 16</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.13.00 06/09
09/28/2016 09:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:54
02 Std. Gas	0.081	09:54
03 Room Air	0.000	09:55
04 Std. Gas	0.082	09:55
05 Room Air	0.000	09:56
06 Std. Gas	0.083	09:56
07 Room Air	0.000	09:57

Lot No = 09216080A3

Cyl No = 44

Exp Date = 05/05/2018

County = 53

Oper No. = 010101



Operator Signature
DARCY ABERLE

Remarks:

Form 106-I8000