



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59281 (03-2014)

Serial Number <b>80-004937</b>	Instrument Location <b>Cass Co. Jail</b>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P). ✓
  - B. ACA Test (Level 1, Function C). ✓
  - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test). ✓ - Detected (x2)
- 7. Repair and/or Maintenance Performed (if Any): None
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date <b>3/23/2016 ✓</b>
Reviewed By (Crime Laboratory Use Only) <b>Deb Shanaver</b>	Date <b>28 Mar 2016</b>

This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota is certified to be a true and correct copy of the documents received.

Certified By:  Certified Date: 01 Nov. 16

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004937  
Location = CASS      8164.13.00 06/09  
03/23/2016      23:43

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:      Enabled  
Start Test Sequence:      DABACABA  
Display Prelim Rslt?      Yes  
Display Third Digit?      Yes  
Inhib Printer(Y/N)?      No  
Display Volume?      No  
Disable On Memfull?      Yes  
# of Print Copies?      1  
Select Std (D/W/I)?      Dry  
Standard Value?      0.080  
Standard Lot #?      29614080A2  
Standard Cyl #?      36  
Standard Expiration?      12/05/2016  
Oper No?      050910

Flow Cal. Date:      01/18/2010  
Slope      730  
Intercept      -779991

IR Calibration Date:      04/08/2015

	3um	9um
0th Coef(*100):	-10967	-18591
1st Coef(*100):	254229	128431
2nd Coef(*100):	2837	1327
H2O adj(mg/l*10k):	348	392

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHAD THOMPSON

Remarks:

Print test


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004937  
Location = CASS      8164.13.00 06/09  
03/23/2016      23:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	23:44
02 Std. Gas	0.081	23:44
03 Room Air	0.000	23:44
04 Std. Gas	0.081	23:45
05 Room Air	0.000	23:45
06 Std. Gas	0.081	23:46
07 Room Air	0.000	23:46

Lot No = 29614080A2  
Cyl No = 36  
Exp Date = 12/05/2016  
County = 09      Oper No. = 050910



Operator Signature  
CHAD THOMPSON

Remarks: ACA Test

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004937  
Location = CASS      8164.13.00 06/09  
03/23/2016      23:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	23:49
02 Room Air	RFI*	23:49

\*Invalid Test  
Inhibited - RFI

Lot No = 29614080A2  
Cyl No = 36  
Exp Date = 12/05/2016  
County = 09      Oper No. = 050910

 3524

Operator Signature  
CHAD THOMPSON

Remarks:

RFI Test  
Detected (x2)

Form 106-I8000