



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <b>80-005950</b>	Instrument Location <b>SHER</b>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): \_\_\_\_\_
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Dustin Hoffman</i>	Date <b>05-24-2022</b>
--	---------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Janelli Portscheller</i>	Certified Date <b>01 June 22</b>
--	-------------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005950  
Location = SHER                    8164.14.00 09/16  
05/24/2022                            11:46

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?


Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            26021080A1  
Standard Cyl #?            34  
Standard Expiration?       10/05/2023  
Oper No?                    048909  
  
Flow Cal. Date:            06/20/2013  
Slope                      713  
Intercept                  -723515

IR Calibration Date:       06/20/2013  
                                    3um                      9um

-----  
0th Coef(\*100):        -37081                      -23816  
1st Coef(\*100):        290549                      133411  
2nd Coef(\*100):        552                         1090  
H2O adj(mg/l\*10k):    683                         502

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
\_\_\_\_\_  
Operator Signature  
DELBERT HOFFMANN

Remarks:

PRINT TEST


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005950  
Location = SHER      8164.14.00 09/16  
05/24/2022      11:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:47
02 Std. Gas	0.081	11:47
03 Room Air	0.000	11:48
04 Std. Gas	0.080	11:48
05 Room Air	0.000	11:49
06 Std. Gas	0.082	11:49
07 Room Air	0.000	11:50

Lot No = 26021080A1  
Cyl No = 34  
Exp Date = 10/05/2023  
County = 42      Oper No. = 048909

  
Operator Signature  
DELBERT HOFFMANN

Remarks:

ACA TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005950  
Location = SHER      8164.14.00 09/16  
05/24/2022      11:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:55
02 Std. Gas	RFI*	11:55
03 Room Air	0.000	11:56

\*Invalid Test  
Inhibited - RFI

Lot No = 26021080A1  
Cyl No = 34  
Exp Date = 10/05/2023  
County = 42      Oper No. = 048909

  
Operator Signature  
DELBERT HOFFMANN

Remarks: RFI TEST