



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6505 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
  - Gas Cylinder  Yes or  No (If Yes, Lot # 26021080A1 Cyl. # 23)
  - Keys  Yes or  No
  - Power Cord  Yes or  No
2.  Download Data
3.  Upload Operator File
4.  Current Location Code: MDHL
5.  Battery Check
  - Was the external battery pack replaced? Yes or  No
6.  O-Rings
  - Replaced Simulator O-Ring Yes or  No
  - Replaced Breath Tube O-Ring Yes or  No

B. General Setup and Checks:

1.  Diagnostics passed and instrument in "Ready" mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone:  CST or  MDT (Time on test records will be in time zone circled).
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G).  
Display: 901 psi Regulator: 950 psi  
Display and Regulator ± 50 psi of each other  Yes or  No  
Gas tank tare necessary? Yes or  No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1.  Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value ≤ 0.03 AC: 0.020 AC  
Sim. Ser #: MP5318  
Lot #: 202201A  
Exp. Date: 11Jan24  
 Results ± 0.005 of known AC

Toxicology Unit - Breath Alcohol Section	Qualtrax ID: 11698 Revision: 1
Intoxilyzer 8000 Annual Inspection	Status: Published Published: 03/24/2023
UNCONTROLLED WHEN PRINTED	Page 1 of 2

2.  Configure simulator for the following test (Level 1,S).  
 Wet Calibration Check - High AC (Level 1,C)  
 Known Value  $\geq$  0.25 AC: 0.300 AC  
 Sim. Ser #: MP3067  
 Lot #: 202201F  
 Exp. Date: 18Jan24  
 Results  $\pm$  5% AC of known AC
3.  Configure dry gas standard for the remaining tests (Level 1,S).  
 Known Value : 0.090 AC  
 Gas Cylinder Lot #: 26021080A1  
 Cylinder #: 3  
 Exp. Date: 10/5/23
4.  Interferent Check (Level 1,B)  
 Known Value: 0.10 AC + 0.05% Acetone  
 Sim. Ser #: DR7345  
 Lot #: 1CS7  
 Exp. Date: NIA  
 Display reads "Interferent Detect"
5.  RFI Check (CMS Mode)  
 Display reads "RFI Detect"
6.  Dry Calibration Check (Level 1,C)  
 Test 1 0.079      Test 4 0.080      Test 7 0.078  
 Test 2 0.079      Test 5 0.079      Test 8 0.079  
 Test 3 0.079      Test 6 0.079      Test 9 0.079  
 Average 0.079  
 Results  $\pm$  0.005 AC of known AC

D. Remarks/Maintenance: NIA

---



---

Instrument is acceptable to be used in the field.  Yes or No

If No, state reason(s) why: \_\_\_\_\_

---

If Yes, change location code back to A.4.

[Signature]  
 Inspector Signature

26 May 2023  
 Date

[Signature]  
 Reviewer

30 May 2023  
 Date

Toxicology Unit - Breath Alcohol Section	Qualtrax ID: 11698 Revision: 1
Intoxilyzer 8000 Annual Inspection	Status: Published Published: 03/24/2023
UNCONTROLLED WHEN PRINTED	Page 2 of 2

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006505  
Location = TOXL                    8164.14.00 09/16  
05/26/2023                            11:20

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            26021080A1  
Standard Cyl #?            23  
Standard Expiration?       10/05/2023  
Oper No?                    133237

Flow Cal. Date:            08/20/2015  
Slope                      671  
Intercept                  -642269

IR Calibration Date:        08/20/2015

	3um	9um
0th Coef(*100):	-29025	-17051
1st Coef(*100):	252705	133345
2nd Coef(*100):	3840	1546
H2O adj(mg/l*10k):	817	461

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
\_\_\_\_\_  
Operator Signature  
ANNA NAREHOOD

Remarks: *Print Test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006505  
Location = TOXL      8164.14.00 09/16  
05/26/2023      11:29

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:30
02 Std. Sol.	0.019	11:30
03 Room Air	0.000	11:31
04 Std. Sol.	0.019	11:32
05 Room Air	0.000	11:32
06 Std. Sol.	0.020	11:33
07 Room Air	0.000	11:34

08 Sim Temp = 34.0°C

Simul Ser No = MP5318  
Std Sol No = 202201A  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: *LOW Ac Check- 0.020Ac*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006505  
Location = TOXL      8164.14.00 09/16  
05/26/2023      11:35

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:35
02 Std. Sol.	0.299	11:36
03 Room Air	0.000	11:37
04 Std. Sol.	0.299	11:38
05 Room Air	0.000	11:39
06 Std. Sol.	0.299	11:40
07 Room Air	0.000	11:40

08 Sim Temp = 34.0°C

Simul Ser No = MP3067

Std Sol No = 202201F

County = 08

Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: High AC Check - 0.300Ac

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-006505  
Location = TOXL                                      8164.14.00 09/16  
05/26/2023    11:41

Test	AC	Time
01 Room Air	0.000	11:42
02 *Subject Test	INT*	11:43
03 Room Air	0.000	11:43

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                                      Weight = NA  
Test = OTH                                      Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 26021080A1  
Cyl No = 3  
Expiration Date = 10/05/2023  
County = 08                                      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Interferent check

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-006505  
Location = TOXL                              8164.14.00 09/16  
05/26/2023    11:44

Test	AC	Time
01 Diagnostic	OK	11:45
02 Room Air	0.000	11:45
03 *Subject Test	RFI*	11:46
04 Room Air	0.000	11:46

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                                      Weight = NA  
Test = OTH    Cit = RFI CHECK  
Dr. Lic. = ND/TES989643  
Lot No = 26021080A1  
Cyl No = 3  
Expiration Date = 10/05/2023  
County = 08    Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006505  
Location = TOXL      8164.14.00 09/16  
05/26/2023      11:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:47
02 Std. Gas	0.079	11:47
03 Room Air	0.000	11:48
04 Std. Gas	0.079	11:48
05 Room Air	0.000	11:49
06 Std. Gas	0.079	11:49
07 Room Air	0.000	11:50

Lot No = 26021080A1  
Cyl No = 3  
Exp Date = 10/05/2023  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #1-3

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006505  
Location = TOXL      8164.14.00 09/16  
05/26/2023      11:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:51
02 Std. Gas	0.080	11:52
03 Room Air	0.000	11:52
04 Std. Gas	0.079	11:52
05 Room Air	0.000	11:53
06 Std. Gas	0.079	11:53
07 Room Air	0.000	11:54

Lot No = 26021080A1  
Cyl No = 3  
Exp Date = 10/05/2023  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006505  
Location = TOXL      8164.14.00 09/16  
05/26/2023      11:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:55
02 Std. Gas	0.078	11:56
03 Room Air	0.000	11:56
04 Std. Gas	0.079	11:56
05 Room Air	0.000	11:57
06 Std. Gas	0.079	11:57
07 Room Air	0.000	11:58

Lot No = 26021080A1  
Cyl No = 3  
Exp Date = 10/05/2023  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA NAREHOOD

Remarks: Dry Cal check #7-9

Form 106-I8000