



**ETHANOL GAS STANDARD CYLINDER REPORT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Sebediah Kohler</i>		
Location <i>Hettinger</i>	Intoxilyzer® Serial Number <i>80-007095</i>	
Gas Lot Number <i>34921080A1</i>	Gas Cylinder Number <i>29</i>	Gas Expiration Date <i>02/05/2024</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).  
 A. 0. *0.09* AC  
 B. 0. *0.02* AC  
 C. 0. *0.02* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:  
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).  
 B. ACA Test Record.

Chemical Test Operator Signature <i>Sebediah Kohler</i>	Date <i>2/16/23</i>
Reviewed By (Crime Laboratory Use Only) <i>AEN</i>	Date <i>27 Feb 2023</i>

*\*confirmed with remote upload  
 that county code has ~~been~~  
 been changed. AEN 16 Mar 23*


Intoxilyzer Test Record and Checklist  
NDCAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007095  
Location = ~~ADAM~~ HETT      8154.16.00 09/18  
02/16/2023      AEN      14:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:28
02 Std. Gas	0.082	14:29
03 Room Air	0.000	14:29
04 Std. Gas	0.082	14:29
05 Room Air	0.000	14:30
06 Std. Gas	0.082	14:30
07 Room Air	0.000	14:31

Lot No = 34921080A1  
Cyl No = 29  
Exp Date = 02/05/2024  
County = 01      Oper No. = 132666

  
\_\_\_\_\_  
Operator Signature  
JEDEDIAH KOHLER

Remarks:

Form 106-I8000