



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Jehediah Kohler</i>		
Location <i>Hettinger</i>	Intoxilyzer® Serial Number <i>80-007095</i>	
Gas Lot Number <i>34921080A1</i>	Gas Cylinder Number <i>29</i>	Gas Expiration Date <i>02/05/2024</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.082* AC
 - B. *0.082* AC
 - C. *0.082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Jehediah Kohler</i>	Date <i>2/16/23</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>27 Feb 2023</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = ADAM 8154.16.00 09/18
02/16/2023 14:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:28
02 Std. Gas	0.082	14:29
03 Room Air	0.000	14:29
04 Std. Gas	0.082	14:29
05 Room Air	0.000	14:30
06 Std. Gas	0.082	14:30
07 Room Air	0.000	14:31

Lot No = 34921080A1
Cyl No = 29
Exp Date = 02/05/2024
County = 01 Oper No. = 132666


Operator Signature
JEDEDIAH KOHLER

Remarks:

Form 106-I8000