



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-007095	Instrument Location HETT
Reason for Install/Repair	
<input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000. **OK**
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): **N/A Field Insp**
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature And Koliba	Date 2/16/23
--	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By [Signature]	Certified Date 27 Feb 2023
---	--------------------------------------

**Confirmed with remote upload that county code has been changed. AEN 16Mar23.*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = ~~ADAM~~ **HETT** 8164.16.00 09/18
02/16/2023 **AEN** 13:56

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34921080A1
Standard Cyl #? 29
Standard Expiration? 02/05/2024
Oper No? 132666

Flow Cal. Date: 09/27/2018
Slope 683
Intercept -637452

IR Calibration Date: 09/27/2018
 3um 9um

0th Coef(*100):	-59629	-23097
1st Coef(*100):	252379	137307
2nd Coef(*100):	2489	1549
H2O adj(mg/l*10k):	1414	594

***** Printer Test End *****


Operator Signature
JEDEDIAH KOHLER

Remarks:

Form 106-18000

Intoxilyzer Test Record and Checklist
NDCAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = ~~ADAM~~ **HETT** 8154.16.00 09/18
02/16/2023 **AEN** 14:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:28
02 Std. Gas	0.082	14:29
03 Room Air	0.000	14:29
04 Std. Gas	0.082	14:29
05 Room Air	0.000	14:30
06 Std. Gas	0.082	14:30
07 Room Air	0.000	14:31

Lot No = 34921080A1
Cyl No = 29
Exp Date = 02/05/2024
County = 01 Oper No. = 132666



Operator Signature
JEDEDIAH KOHLER

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = ~~ADAM~~ HETT 8164.16.00 09/18
02/16/2023 AEN 14:33

Test	AC	Time
01 Room Air	RFI*	14:35
02 Room Air	0.000	14:35

*Invalid Test
Inhibited - RFI

Sub Name = KOHLER, JEDEDIAH SCOTT
Sub DOB = 08/20/1987
Sub Sex = Male Weight = NA
Test = OTH Cit = NA
Dr. Lic. = ND/KOH872036
Lot No = 34921080A1
Cyl No = 29
Expiration Date = 02/05/2024
County = 01 Oper No. = 132666



Operator Signature
JEDEDIAH KOHLER

Remarks:

Form 106-18000