





ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (08/2018)

Chemical Test Operator Name (Print) HALL, ALLEN J		
Location KENMARE POLICE DEPT	Intoxilyzer® Serial Number 80-006691	
Gas Lot Number 01923080A3	Gas Cylinder Number 61	Gas Expiration Date 2-17-25

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
A. **0.081** AC
B. **0.081** AC
C. **0.081** AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature  A HALL	Date 2-27-23
Reviewed By (Crime Laboratory Use Only) 	Date 28 Feb 2023

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = KMPD 8164.14.00 09/16
02/27/2023 14:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:59
02 Std. Gas	0.081	14:59
03 Room Air	0.000	15:00
04 Std. Gas	0.081	15:00
05 Room Air	0.000	15:01
06 Std. Gas	0.081	15:01
07 Room Air	0.000	15:02

Lot No = 01923080A3
Cyl No = 61
Exp Date = 02/05/2025
County = 51 Oper No. = 133223



Operator Signature
ALLEN HALL

Remarks:

Form 106-I8000