



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Trevor K. Bergerson</i>		
Location <i>BOPD</i>	Intoxilyzer® Serial Number <i>80-007093</i>	
Gas Lot Number <i>34921080A1</i>	Gas Cylinder Number <i>044</i>	Gas Expiration Date <i>02/05/2024</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
A. 0. *080* AC
B. 0. *080* AC
C. 0. *080* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
B. ACA Test Record.

Chemical Test Operator Signature <i>Trevor K. Bergerson</i>	Date <i>3/3/2023</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>06Mar2023</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007093
Location = BOPD 8164.16.00 09/18
03/03/2023 23:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	23:35
02 Std. Gas	0.080	23:36
03 Room Air	0.000	23:36
04 Std. Gas	0.080	23:36
05 Room Air	0.000	23:37
06 Std. Gas	0.080	23:37
07 Room Air	0.000	23:38

Lot No = 34921080A1
Cyl No = 44
Exp Date = 02/05/2024
County = 06 Oper No. = 130538


Operator Signature
TREVOR BERGERSON

Remarks:

Form 106-I8000