



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Jonathan Lampi</i>		
Location <i>GFPD</i>	Intoxilyzer® Serial Number <i>80-006678</i>	
Gas Lot Number <i>01923080A3</i>	Gas Cylinder Number <i>57</i>	Gas Expiration Date <i>02-05-2025</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0083* AC
 - B. *0083* AC
 - C. *0083* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Jonathan Lampi</i>	Date <i>03-06-23</i>
Reviewed By (Crime Laboratory Use Only) <i>Chris Muel</i>	Date <i>13Mar2023</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006678
Location = ~~TOX~~ **GFPD** 8164.14.00 09/16
03/06/2023 12:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:43
02 Std. Gas	0.083	12:43
03 Room Air	0.000	12:44
04 Std. Gas	0.083	12:44
05 Room Air	0.000	12:44
06 Std. Gas	0.083	12:45
07 Room Air	0.000	12:45

Lot No = 01923080A3
Cyl No = 57
Exp Date = 02/05/2025
County = 18 Oper No. = 039947

 039947
Operator Signature
JONATHAN LAMPI

Remarks:

Ethanol Gasstandard & Cylinder Report
Form 106-I8000

-Talked to Jonathan Lampi and connected to Intox
to ensure correct location code.
AEN 13Mar2023