



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>La Badda, Seth</i>		
Location <i>Minot Police Department</i>	<i>515 2nd Ave SW Minot, ND 58701</i>	Intoxilyzer® Serial Number <i>80-006503</i>
Gas Lot Number <i>34921080A1</i>	Gas Cylinder Number <i>31</i>	Gas Expiration Date <i>02/05/2024</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits. Ex. 0.081).
 A. *0.082* AC
 B. *0.082* AC
 C. *0.082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 B. ACA Test Record.

Chemical Test Operator Signature <i>Seth LaBadda</i> 436/8125	Date <i>March 7, 2023</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>09Mar2023</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = MIPD 8164.14.00 09/16
03/07/2023 16:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:53
02 Std. Gas	0.082	16:54
03 Room Air	0.000	16:54
04 Std. Gas	0.082	16:55
05 Room Air	0.000	16:55
06 Std. Gas	0.082	16:56
07 Room Air	0.000	16:56

Lot No = 34921080A1
Cyl No = 31
Exp Date = 02/05/2024
County = 51 Oper No. = 132527



436/8105

Operator Signature
SETH LABODDA

Remarks:

Form 106-I8000