



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

| | |
|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------|
| Serial Number <i>80-006506</i> | Instrument Location <i>Ramsey Co #2</i> |
| Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____ | |

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000. *N/A - working on getting phone line installed*
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): *N/A*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency. *N/A - first install of instrument*
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

| | |
|-------------------------------------------------|---------------------------|
| Field Inspector Signature <i>[Signature]</i> | Date <i>03/20/2023</i> |
|-------------------------------------------------|---------------------------|

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

| | |
|---------------------------------------------|------------------------------------|
| Reviewed/Certified By <i>[Signature]</i> | Certified Date <i>21Mar2023</i> |
|---------------------------------------------|------------------------------------|

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = DEVL 8164.14.00 09/16
03/20/2023 14:02

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34921080A1
Standard Cyl #? 27
Standard Expiration? 02/05/2024
Oper No? 132008

Flow Cal. Date: 08/20/2015
Slope 657
Intercept -627881

IR Calibration Date: 08/20/2015

| | 3um | 9um |
|--------------------|--------|--------|
| 0th Coef(*100): | -22470 | -18919 |
| 1st Coef(*100): | 266505 | 136109 |
| 2nd Coef(*100): | 3702 | 1644 |
| H2O adj(mg/l*10k): | 709 | 510 |

***** Printer Test End *****



Operator Signature
WILLIAM EARL

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = DEVL 8164.14.00 09/16
03/20/2023 14:02

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 14:03 |
| 02 Std. Gas | 0.082 | 14:03 |
| 03 Room Air | 0.000 | 14:04 |
| 04 Std. Gas | 0.082 | 14:04 |
| 05 Room Air | 0.000 | 14:05 |
| 06 Std. Gas | 0.081 | 14:05 |
| 07 Room Air | 0.000 | 14:05 |

Lot No = 34921080A1
Cyl No = 27
Exp Date = 02/05/2024
County = 36 Oper No. = 132008



Operator Signature
WILLIAM EARL

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = DEVL 8164.14.00 09/16
03/20/2023 14:06

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 14:07 |
| 02 Std. Gas | 0.081 | 14:07 |
| 03 Room Air | RFI* | 14:07 |
| 04 Room Air | 0.000 | 14:08 |

*Invalid Test
Inhibited - RFI

Lot No = 34921080A1
Cyl No = 27
Exp Date = 02/05/2024
County = 36 Oper No. = 132008



Operator Signature
WILLIAM EARL

Remarks:

Form 106-I8000