



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 3059 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # _____ Cyl. # _____)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: TOXL
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
Display: 843 psi Regulator: 850 psi
Display and Regulator ± 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP5318
Lot #: 202201A
Exp. Date: 1Jan24
 Results ± 0.005 of known AC

Toxicology Unit - Breath Alcohol Section	Qualtrax ID: 11698 Revision: 1
Intoxilyzer 8000 Annual Inspection	Status: Published Published: 03/24/2023
UNCONTROLLED WHEN PRINTED	Page 1 of 2

2. Configure simulator for the following test (Level 1,S).
 Wet Calibration Check - High AC (Level 1,C)
 Known Value \geq 0.25 AC: 0.300 AC
 Sim. Ser #: MPU039
 Lot #: 262201F
 Exp. Date: 18Jan24
 Results \pm 5% AC of known AC
3. Configure dry gas standard for the remaining tests (Level 1,S).
 Known Value : 0.080 AC
 Gas Cylinder Lot #: 26021080A1
 Cylinder #: 25
 Exp. Date: 10/5/23
4. Interferent Check (Level 1,B)
 Known Value: 0.10 AC + 0.05% Acetone
 Sim. Ser #: DR3841
 Lot #: ICS 7
 Exp. Date: NIA
 Display reads "Interferent Detect"
5. RFI Check (CMS Mode)
 Display reads "RFI Detect"
6. Dry Calibration Check (Level 1,C)
 Test 1 0.082 Test 4 0.081 Test 7 0.081
 Test 2 0.082 Test 5 0.082 Test 8 0.081
 Test 3 0.081 Test 6 0.081 Test 9 0.082
 Average 0.081
 Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: NIA

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4.

[Signature]
 Inspector Signature

04 Apr 2023
 Date

Janelle Pertschiller
 Reviewer

16 Apr 2023
 Date

Toxicology Unit - Breath Alcohol Section	Qualtrax ID: 11698 Revision: 1
Intoxilyzer 8000 Annual Inspection	Status: Published Published: 03/24/2023
UNCONTROLLED WHEN PRINTED	Page 2 of 2

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.14.00 09/16
04/06/2023 11:40

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&* () _ + ?

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&* () _ + ?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34917080A3
Standard Cyl #? 35
Standard Expiration? 02/05/2020
Oper No? 133237

Flow Cal. Date: 11/28/2018
Slope 589
Intercept -276335

IR Calibration Date: 12/04/2018

	3um	9um
0th Coef(*100):	-25330	-15825
1st Coef(*100):	278286	137135
2nd Coef(*100):	1992	1257
H2O adj(mg/l*10k):	525	321

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: Print Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.14.00 09/16
04/06/2023 11:42

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:42
02 Std. Sol.	0.018	11:43
03 Room Air	0.000	11:43
04 Std. Sol.	0.019	11:44
05 Room Air	0.000	11:45
06 Std. Sol.	0.020	11:45
07 Room Air	0.000	11:46

08 Sim Temp = 34.0°C

Simul Ser No = MP5318
Std Sol No = 202201A
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *LOW AC CHECK-0.020AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.14.00 09/16
04/06/2023 11:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:47
02 Std. Sol.	0.301	11:48
03 Room Air	0.000	11:48
04 Std. Sol.	0.304	11:49
05 Room Air	0.000	11:50
06 Std. Sol.	0.303	11:50
07 Room Air	0.000	11:51

08 Sim Temp = 34.0°C

Simul Ser No = MP6039

Std Sol No = 202201F

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks:

High Ac Check - 0.300Ac

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.14.00 09/16
04/06/2023 11:53

Test	AC	Time
01 Room Air	0.000	11:54
02 *Subject Test	INT*	11:54
03 Room Air	0.000	11:55

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 26021080A1
Cyl No = 25
Expiration Date = 10/05/2023
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks:

Interferent Check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.14.00 09/16
04/06/2023 11:56

Test	AC	Time
01 Diagnostic	OK	11:57
02 Room Air	0.000	11:57
03 *Subject Test	RFI*	11:57
04 Room Air	0.000	11:58

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 26021080A1
Cyl No = 25
Expiration Date = 10/05/2023
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI Check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.14.00 09/16
04/06/2023 11:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:58
02 Std. Gas	0.082	11:59
03 Room Air	0.000	11:59
04 Std. Gas	0.082	12:00
05 Room Air	0.000	12:00
06 Std. Gas	0.081	12:01
07 Room Air	0.000	12:01

Lot No = 26021080A1
Cyl No = 25
Exp Date = 10/05/2023
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks:

Dry cal check #1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.14.00 09/16
04/06/2023 12:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:02
02 Std. Gas	0.081	12:02
03 Room Air	0.000	12:03
04 Std. Gas	0.082	12:03
05 Room Air	0.000	12:03
06 Std. Gas	0.081	12:04
07 Room Air	0.000	12:04

Lot No = 26021080A1
Cyl No = 25
Exp Date = 10/05/2023
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #4-4

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.14.00 09/16
04/06/2023 12:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:05
02 Std. Gas	0.081	12:06
03 Room Air	0.000	12:06
04 Std. Gas	0.081	12:07
05 Room Air	0.000	12:07
06 Std. Gas	0.082	12:08
07 Room Air	0.000	12:08

Lot No = 26021080A1
Cyl No = 25
Exp Date = 10/05/2023
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks:

Dry cal check #7-9

Form 106-I8000