



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 4198 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
  - Gas Cylinder Yes or No (If Yes, Lot # \_\_\_\_\_ Cyl. # \_\_\_\_\_)
  - Keys Yes or No
  - Power Cord Yes or No
2.  Download Data
3.  Upload Operator File
4.  Current Location Code: TOXL
5.  Battery Check
  - Was the external battery pack replaced? Yes or No
6.  O-Rings
  - Replaced Simulator O-Ring Yes or No
  - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1.  Diagnostics passed and instrument in "Ready" mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G).  
Display: 688 psi Regulator: 700 psi  
Display and Regulator ± 50 psi of each other Yes or No  
Gas tank tare necessary? Yes or No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1.  Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value ≤ 0.03 AC: 0.020 AC  
Sim. Ser #: MP5318  
Lot #: 202201A  
Exp. Date: 11Jan24  
 Results ± 0.005 of known AC

Toxicology Unit - Breath Alcohol Section	Qualtrax ID: 11698 Revision: 1
Intoxilyzer 8000 Annual Inspection	Status: Published Published: 03/24/2023
UNCONTROLLED WHEN PRINTED	Page 1 of 2



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004198  
Location = TOXL                    8164.14.00 09/16  
04/21/2023                            09:46

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:                    Enabled  
Start Test Sequence:                DABACABA  
Display Prelim Rslt?                Yes  
Display Third Digit?                Yes  
Inhib Printer(Y/N)?                No  
Display Volume?                    No  
Disable On Memfull?                Yes  
# of Print Copies?                  1  
Select Std (D/W/I)?                Dry  
Standard Value?                    0.080  
Standard Lot #?                    26021080A1  
Standard Cyl #?                    25  
Standard Expiration?                10/05/2023  
Oper No?                            133237

Flow Cal. Date:                    01/09/2012  
Slope                                637  
Intercept                            -467208

IR Calibration Date:                05/18/2016  
    3um                                9um

-----  
0th Coef(\*100):                    -19233                                -23050  
1st Coef(\*100):                    267031                                132119  
2nd Coef(\*100):                    2449                                    1142  
H2O adj(mg/l\*10k):                329                                    398

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
ANNA NAREHOOD

Remarks:    Print Test

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004198  
Location = TOXL      8164.14.00 09/16  
04/21/2023      09:47

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:48
02 Std. Sol.	0.020	09:50
03 Room Air	0.000	09:51
04 Std. Sol.	0.020	09:52
05 Room Air	0.000	09:52
06 Std. Sol.	0.021	09:53
07 Room Air	0.000	09:53

08 Sim Temp = 34.0°C

Simul Ser No = MP5318  
Std Sol No = 202201A  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: *Low AC Check - 0.020AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004198  
Location = TOXL      8164.14.00 09/16  
04/21/2023      09:54

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:55
02 Std. Sol.	0.300	09:56
03 Room Air	0.000	09:56
04 Std. Sol.	0.305	09:57
05 Room Air	0.000	09:58
06 Std. Sol.	0.305	09:58
07 Room Air	0.000	09:59

08 Sim Temp = 34.0°C

Simul Ser No = MP6039  
Std Sol No = 202201F  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: High AC Check - 0.300AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-004198  
Location = TOXL                              8164.14.00 09/16  
04/21/2023    10:01

Test	AC	Time
01 Room Air	0.000	10:02
02 *Subject Test	INT*	10:02
03 Room Air	0.000	10:03

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                              Weight = NA  
Test = OTH                              Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 26021080A1  
Cyl No = 25  
Expiration Date = 10/05/2023  
County = 08                              Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Interferent Check

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004198  
Location = TOXL      8164.14.00 09/16  
04/21/2023      10:03

Test	AC	Time
01 Diagnostic	OK	10:04
02 Room Air	0.000	10:05
03 *Subject Test	RFI*	10:05
04 Room Air	0.000	10:05

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = RFI CHECK  
Dr. Lic. = ND/TES989643  
Lot No = 26021080A1  
Cyl No = 25  
Expiration Date = 10/05/2023  
County = 08      Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

  
\_\_\_\_\_  
Operator Signature  
ANNA NAREHOOD

Remarks: RFI Check

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004198  
Location = TOXL      8164.14.00 09/16  
04/21/2023      10:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:06
02 Std. Gas	0.079	10:06
03 Room Air	0.000	10:07
04 Std. Gas	0.079	10:07
05 Room Air	0.000	10:08
06 Std. Gas	0.080	10:08
07 Room Air	0.000	10:09

Lot No = 26021080A1  
Cyl No = 25  
Exp Date = 10/05/2023  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA NAREHOOD

Remarks: Dry Cal Check #1-3

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004198  
Location = TOXL      8164.14.00 09/16  
04/21/2023      10:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:09
02 Std. Gas	0.079	10:10
03 Room Air	0.000	10:10
04 Std. Gas	0.079	10:11
05 Room Air	0.000	10:11
06 Std. Gas	0.079	10:12
07 Room Air	0.000	10:12

Lot No = 26021080A1  
Cyl No = 25  
Exp Date = 10/05/2023  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA NAREHOOD

Remarks: DRY CAL CHECK #4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004198  
Location = TOXL      8164.14.00 09/16  
04/21/2023      10:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:13
02 Std. Gas	0.080	10:13
03 Room Air	0.000	10:14
04 Std. Gas	0.080	10:14
05 Room Air	0.000	10:15
06 Std. Gas	0.079	10:15
07 Room Air	0.000	10:15

Lot No = 26021080A1  
Cyl No = 25  
Exp Date = 10/05/2023  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check # 7-9

Form 106-I8000