



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 5957 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # 26021080A1 Cyl. # 4)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: MAND
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
Display: 788 psi Regulator: 800 psi
Display and Regulator \pm 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: MP5318
Lot #: 202201A
Exp. Date: 11Jan24
 Results \pm 0.005 of known AC

Toxicology Unit - Breath Alcohol Section	Qualtrax ID: 11698 Revision: 1
Intoxilyzer 8000 Annual Inspection	Status: Published Published: 03/24/2023
UNCONTROLLED WHEN PRINTED	Page 1 of 2

2. Configure simulator for the following test (Level 1,S).
 Wet Calibration Check - High AC (Level 1,C)
 Known Value \geq 0.25 AC: 0.300 AC
 Sim. Ser #: MP6039
 Lot #: 202201F
 Exp. Date: 18 Jan 24
 Results \pm 5% AC of known AC
3. Configure dry gas standard for the remaining tests (Level 1,S).
 Known Value : 0.080 AC
 Gas Cylinder Lot #: 34921080A1
 Cylinder #: 49
 Exp. Date: 2/5/24
4. Interferent Check (Level 1,B)
 Known Value: 0.10 AC + 0.05% Acetone
 Sim. Ser #: DR3841
 Lot #: ICS 7
 Exp. Date: N/A
 Display reads "Interferent Detect"
5. RFI Check (CMS Mode)
 Display reads "RFI Detect"
6. Dry Calibration Check (Level 1,C)
 Test 1 0.080 Test 4 0.081 Test 7 0.080
 Test 2 0.080 Test 5 0.081 Test 8 0.080
 Test 3 0.079 Test 6 0.080 Test 9 0.081
 Average 0.080
 Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4.

Au Mill
 Inspector Signature

24 Apr 2023
 Date

Janelle Pertscheller
 Reviewer

25 Apr 2023
 Date

Toxicology Unit - Breath Alcohol Section	Qualtrax ID: 11698 Revision: 1
Intoxilyzer 8000 Annual Inspection	Status: Published Published: 03/24/2023
UNCONTROLLED WHEN PRINTED	Page 2 of 2

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = TOXL 8164.14.00 09/16
04/24/2023 11:44

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 26021080A1
Standard Cyl #? 4
Standard Expiration? 10/05/2023
Oper No? 133237

Flow Cal. Date: 06/21/2013
Slope 647
Intercept -613317

IR Calibration Date: 05/11/2015
 3um 9um

0th Coef(*100): -16666 -18045
1st Coef(*100): 278478 132949
2nd Coef(*100): 3303 1397
H2O adj(mg/l*10k): 379 351

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: *Print Test*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = TOXL 8164.14.00 09/16
04/24/2023 11:47

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:47
02 Std. Sol.	0.020	11:48
03 Room Air	0.000	11:49
04 Std. Sol.	0.020	11:49
05 Room Air	0.000	11:50
06 Std. Sol.	0.019	11:51
07 Room Air	0.000	11:51


08 Sim Temp = 34.0°C

Simul Ser No = MP5318

Std Sol No = 202201A

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks:

LOW AC CHECK -0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = TOXL 8164.14.00 09/16
04/24/2023 11:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:53
02 Std. Sol.	0.297	11:54
03 Room Air	0.000	11:55
04 Std. Sol.	0.298	11:55
05 Room Air	0.000	11:56
06 Std. Sol.	0.298	11:57
07 Room Air	0.000	11:57

08 Sim Temp = 34.0°C

Simul Ser No = MP6039
Std Sol No = 202201F
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC Check - 0.300AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = TOXL 8164.14.00 09/16
04/24/2023 11:58

Test	AC	Time
01 Room Air	0.000	11:59
02 *Subject Test	INT*	12:00
03 Room Air	0.000	12:00

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 34921080A1
Cyl No = 49
Expiration Date = 02/05/2024
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Interferent check*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = TOXL 8164.14.00 09/16
04/24/2023 12:01

Test	AC	Time
01 Diagnostic	OK	12:01
02 Room Air	0.000	12:02
03 *Subject Test	RFI*	12:04
04 Room Air	0.000	12:05

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 34921080A1
Cyl No = 49
Expiration Date = 02/05/2024
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = TOXL 8164.14.00 09/16
04/24/2023 12:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:07
02 Std. Gas	0.080	12:07
03 Room Air	0.000	12:07
04 Std. Gas	0.080	12:08
05 Room Air	0.000	12:08
06 Std. Gas	0.079	12:09
07 Room Air	0.000	12:09

Lot No = 34921080A1
Cyl No = 49
Exp Date = 02/05/2024
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #1-3

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = TOXL 8164.14.00 09/16
04/24/2023 12:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:10
02 Std. Gas	0.081	12:10
03 Room Air	0.000	12:11
04 Std. Gas	0.081	12:11
05 Room Air	0.000	12:12
06 Std. Gas	0.080	12:12
07 Room Air	0.000	12:13

Lot No = 34921080A1
Cyl No = 49
Exp Date = 02/05/2024
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = TOXL 8164.14.00 09/16
04/24/2023 12:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:14
02 Std. Gas	0.080	12:14
03 Room Air	0.000	12:14
04 Std. Gas	0.080	12:15
05 Room Air	0.000	12:15
06 Std. Gas	0.081	12:16
07 Room Air	0.000	12:16

Lot No = 34921080A1
Cyl No = 49
Exp Date = 02/05/2024
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High Ac Check # 7-a

Form 106-I8000