



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <i>80-003072</i>	Instrument Location <i>STEB</i>
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <i>Annual inspection at crime laboratory</i>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): \_\_\_\_\_
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Wendy Kelly</i>	Date <i>4/24/2023</i>
---	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Ann Hill</i>	Certified Date <i>25Apr2023</i>
--	------------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003072  
Location = STEE                 8164.14.00 09/16  
04/24/2023                        19:34

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%&^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%&^&\*()\_+?

Current Instrument Setup


Data Entry Mode:                Enabled  
Start Test Sequence:            DABACABA  
Display Prelim Rslt?            Yes  
Display Third Digit?            Yes  
Inhib Printer(Y/N)?            No  
Display Volume?                No  
Disable On Memfull?            Yes  
# of Print Copies?              1  
Select Std (D/W/I)?            Dry  
Standard Value?                0.080  
Standard Lot #?                01923080A3  
Standard Cyl #?                60  
Standard Expiration?            02/25/2025  
Oper No?                        131578

Flow Cal. Date:                03/09/2011  
Slope                            682  
Intercept                       -620512

IR Calibration Date:            03/09/2011  
                                  3um                            9um

-----  
0th Coef(\*100):            -24766                        -17499  
1st Coef(\*100):            264559                        138630  
2nd Coef(\*100):            1876                            1114  
H2O adj(mg/l\*10k):        497                            331

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
\_\_\_\_\_  
Operator Signature  
WENDLIN ROHRICH

Remarks:


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = STEE      8164.14.00 09/16  
04/24/2023      19:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:55
02 Std. Gas	0.082	19:55
03 Room Air	0.000	19:56
04 Std. Gas	0.082	19:56
05 Room Air	0.000	19:56
06 Std. Gas	0.081	19:57
07 Room Air	0.000	19:57

Lot No = 01923080A3  
Cyl No = 60  
Exp Date = 02/25/2025  
County = 22      Oper No. = 131578



Operator Signature  
WENDLIN ROHRICH

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = STEE      8164.14.00 09/16  
04/24/2023      19:58

Test	AC	Time
01 Room Air	0.000	19:59
02 *Subject Test	RFI*	19:59
03 Room Air	0.000	20:00

\*Invalid Test  
Inhibited - RFI

Sub Name = K, L NA  
Sub DOB = 09/23/1998  
Sub Sex = Male      Weight = 140  
Test = DUI      Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 01923080A3  
Cyl No = 60  
Expiration Date = 02/25/2025  
County = 22      Oper No. = 131578



Operator Signature  
WENDLIN ROHRICH

Remarks:

Form 106-I8000