



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 4676

Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
  - Gas Cylinder  Yes or No (If Yes, Lot # 26021080A1 Cyl. # 16 ) -
  - Keys  Yes or  No
  - Power Cord  Yes or  No
2.  Download Data
3.  Upload Operator File
4.  Current Location Code: MOSO
5.  Battery Check
  - Was the external battery pack replaced?  Yes or No
6.  O-Rings
  - Replaced Simulator O-Ring  Yes or No
  - Replaced Breath Tube O-Ring  Yes or  No

B. General Setup and Checks:

1.  Diagnostics passed and instrument in "Ready" mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
  - Time Zone:  CST or MDT (Time on test records will be in time zone circled).
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G).
  - Display: 410 psi Regulator: 400 psi
  - Display and Regulator ± 50 psi of each other  Yes or No
  - Gas tank tare necessary?  Yes or  No
  - If Yes, display readings after tare (Level 3,M,C,G):
  - Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1.  Configure simulator for the following test (Level 1,S).
  - Wet Calibration Check - Low AC (Level 1,C)
  - Known Value ≤ 0.03 AC: 0.020 AC
  - Sim. Ser #: MP5318
  - Lot #: 202201A
  - Exp. Date: 11Jan24
  - Results ± 0.005 of known AC

Toxicology Unit - Breath Alcohol Section	Qualtrax ID: 11698 Revision: 1
Intoxilyzer 8000 Annual Inspection	Status: Published Published: 03/24/2023
UNCONTROLLED WHEN PRINTED	Page 1 of 2



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006676  
Location = TOXL                    8164.14.00 09/16  
04/25/2023                            11:05

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer (Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            26021080A1  
Standard Cyl #?            016  
Standard Expiration?        10/05/2023  
Oper No?                    133237

Flow Cal. Date:            08/22/2016  
Slope                        710  
Intercept                   -713681

IR Calibration Date:        08/22/2016  
                                  3um                            9um

-----  
0th Coef(\*100):            -16667                        -22327  
1st Coef(\*100):            259190                        132291  
2nd Coef(\*100):            3393                            1374  
H2O adj (mg/l\*10k):        611                            546

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
ANNA NAREHOOD

Remarks: *Print Test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = TOXL      8164.14.00 09/16  
04/25/2023      11:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:07
02 Std. Sol.	0.017	11:08
03 Room Air	0.000	11:08
04 Std. Sol.	0.018	11:09
05 Room Air	0.000	11:09
06 Std. Sol.	0.020	11:10
07 Room Air	0.000	11:11

08 Sim Temp = 34.0°C

Simul Ser No = MP5318  
Std Sol No = 202201A  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: *LOW AC CHECK - 0.020AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006676  
Location = TOXL                    8164.14.00 09/16  
04/25/2023                            11:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:12
02 Std. Sol.	0.293	11:13
03 Room Air	0.000	11:13
04 Std. Sol.	0.301	11:14
05 Room Air	0.000	11:14
06 Std. Sol.	0.302	11:15
07 Room Air	0.000	11:16

08 Sim Temp = 34.0°C

Simul Ser No = MP6039  
Std Sol No = 202201F  
County = 08                            Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: High AC Check - 0.300Ac

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                    Alcohol Analyzer  
North Dakota Model 8000                SN 80-006676  
Location = TOXL                            8164.14.00 09/16  
04/25/2023                                        11:18

Test	AC	Time
01 Room Air	0.000	11:19
02 *Subject Test	INT*	11:19
03 Room Air	0.000	11:20

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                                Weight = NA  
Test = OTH                                    Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 26021080A1  
Cyl No = 25  
Expiration Date = 10/05/2023  
County = 08                                    Oper No. = 133237



---

Operator Signature  
ANNA NAREHOOD

Remarks: *Interferent check*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006676  
Location = TOXL                    8164.14.00 09/16  
04/25/2023                            11:20

Test	AC	Time
01 Diagnostic	OK	11:21
02 Room Air	0.000	11:21
03 *Subject Test	RFI*	11:21
04 Room Air	0.000	11:22

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                            Weight = NA  
Test = OTH                                Cit = RFI CHECK  
Dr. Lic. = ND/TES989643  
Lot No = 26021080A1  
Cyl No = 25  
Expiration Date = 10/05/2023  
County = 08                                Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
ANNA NAREHOOD

Remarks: RFI CHECK

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = TOXL      8164.14.00 09/16  
04/25/2023      11:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:23
02 Std. Gas	0.083	11:23
03 Room Air	0.000	11:23
04 Std. Gas	0.082	11:24
05 Room Air	0.000	11:24
06 Std. Gas	0.082	11:25
07 Room Air	0.000	11:25

Lot No = 26021080A1  
Cyl No = 25  
Exp Date = 10/05/2023  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #1-3

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = TOXL      8164.14.00 09/16  
04/25/2023      11:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:26
02 Std. Gas	0.082	11:27
03 Room Air	0.000	11:27
04 Std. Gas	0.082	11:27
05 Room Air	0.000	11:28
06 Std. Gas	0.082	11:28
07 Room Air	0.000	11:29

Lot No = 26021080A1  
Cyl No = 25  
Exp Date = 10/05/2023  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks:

Dry Cal check #4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006676  
Location = TOXL                    8164.14.00 09/16  
04/25/2023                            11:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:29
02 Std. Gas	0.083	11:30
03 Room Air	0.000	11:30
04 Std. Gas	0.082	11:31
05 Room Air	0.000	11:31
06 Std. Gas	0.082	11:32
07 Room Air	0.000	11:32

Lot No = 26021080A1  
Cyl No = 25  
Exp Date = 10/05/2023  
County = 08                            Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #7-9

Form 106-I8000