



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <i>00-005362</i>	Instrument Location <i>1201 12th St. SE, Suite C. Watford City, ND</i>
Reason for Install/Repair	
<input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u><i>Annual Inspection at Crime Laboratory</i></u>	

*5005A*

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): \_\_\_\_\_
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date <i>4/27/23</i>
---	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>[Signature]</i>	Certified Date <i>08 May 2023</i>
---	--------------------------------------

*\* Spoke with Kari Stuart. Print test ran on different day b/c forgot to put into 'Ready Mode'. Went back later to finish install.  
-AEN 08 May 2023*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005362  
Location = MCSO                8164.14.00 09/16  
04/27/2023                        14:46

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            33321080A4  
Standard Cyl #?            12  
Standard Expiration?       01/05/2024  
Oper No?                    132007

Flow Cal. Date:            05/22/2020  
Slope                        659  
Intercept                  -620990

IR Calibration Date:        05/26/2020

	3um	9um
0th Coef(*100):	-20883	-25370
1st Coef(*100):	265508	144960
2nd Coef(*100):	936	714
H2O adj(mg/l*10k):	391	379

\*\*\*\*\* Printer Test End \*\*\*\*\*


  
\_\_\_\_\_  
Operator Signature  
KARI STUART

Remarks:

*PRINT TEST*  
Form 106-I8000

STATE OF NORTH DAKOTA }  
COUNTY OF MCKENZIE }<sup>ss</sup>

I certify that the foregoing is a true and correct  
copy of the original document on file in this office.  
Dated at Watford City, North Dakota.

on this 27th day of April, 2023  
 #32110  
Mckenzie County Sheriff's Department



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-005362  
Location = MCSO                                      8164.14.00 09/16  
05/01/2023    12:03

Test	AC	Time
01 Room Air	RFI*	12:04
02 Room Air	RFI*	12:04

\*Invalid Test  
Inhibited - RFI

Sub Name = STUART, KARI ANN MARIE  
Sub DOB = 09/09/1990  
Sub Sex = Female                                      Weight = 150  
Test = DUI    Cit = 2700000000  
Dr. Lic. = ND/STU900000  
Lot No = 33321080A4  
Cyl No = 12  
Expiration Date = 01/05/2024  
County = 27    Oper No. = 132007

KARI STUART #3410  
Operator Signature  
KARI STUART

Remarks: RFI TEST

Form 106-I8000

STATE OF NORTH DAKOTA }  
COUNTY OF MCKENZIE }<sup>ss</sup>

I certify that the foregoing is a true and correct  
copy of the original document on file in this office.  
Dated at Watford City, North Dakota.

on this 10<sup>th</sup> day of May, 2023  
KARI STUART #3410  
McKenzie County Sheriff's Department