



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <u>SN80-004954</u>	Instrument Location <u>WCPD</u>
Reason for Install/Repair	
<input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>Annual Inspection at crime Laboratory</u>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <u>Sasha C McGovern #1160</u>	Date <u>4/27/23</u>
---	---------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <u>Ann Hill</u>	Certified Date <u>27Apr2023</u>
---------------------------------------	---------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004954  
Location = WCPD                    8164.14.00 09/16  
04/27/2023                            04:17

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            33321080A4  
Standard Cyl #?            11  
Standard Expiration?       01/05/2024  
Oper No?                    131581

Flow Cal. Date:            08/27/2012  
Slope                        670  
Intercept                   -570610

IR Calibration Date:        05/31/2017

	3um	9um
0th Coef(*100):	-18765	-20238
1st Coef(*100):	256075	130805
2nd Coef(*100):	2112	1219
H2O adj(mg/l*10k):	442	407

\*\*\*\*\* Printer Test End \*\*\*\*\*

 #1160  
-----  
Operator Signature  
JOSHUA MCGOVEN

Remarks:


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004954  
Location = WCPD      8164.14.00 09/16  
04/27/2023      04:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	04:18
02 Std. Gas	0.081	04:19
03 Room Air	0.000	04:19
04 Std. Gas	0.082	04:20
05 Room Air	0.000	04:20
06 Std. Gas	0.082	04:20
07 Room Air	0.000	04:21

Lot No = 33321080A4  
Cyl No = 11  
Exp Date = 01/05/2024  
County = 27      Oper No. = 131581

  
Operator Signature  
JOSHUA MCGOVEN

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-004954  
Location = WCPD                              8164.14.00 09/16  
04/27/2023    04:22

Test	AC	Time
01 Room Air	0.000	04:25
02 *Subject Test	RFI*	04:25
03 Room Air	0.000	04:26

\*Invalid Test  
Inhibited - RFI

Sub Name = RFI TEST, NA INSTALL  
Sub DOB = 10/10/2000  
Sub Sex = Unknown                              Weight = NA  
Test = OTH    Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 33321080A4  
Cyl No = 11  
Expiration Date = 01/05/2024  
County = 27    Oper No. = 131581

  
Operator Signature  
JOSHUA MCGOVEN

Remarks:

Form 106-I8000