



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6501 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # 34921080A1 Cyl. # 43)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: MNAF
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
Display: 800 psi Regulator: 900 psi
Display and Regulator ± 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP5318
Lot #: 202201A
Exp. Date: 11Jan24
 Results ± 0.005 of known AC

Toxicology Unit - Breath Alcohol Section	Qualtrax ID: 11698 Revision: 1
Intoxilyzer 8000 Annual Inspection	Status: Published Published: 03/24/2023
UNCONTROLLED WHEN PRINTED	Page 1 of 2

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
04/28/2023 10:51

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34921080A1
Standard Cyl #? 43
Standard Expiration? 02/05/2024
Oper No? 133237


Flow Cal. Date: 06/13/2019
Slope 651
Intercept -777821

IR Calibration Date: 06/13/2019

	3um	9um
--	-----	-----

0th Coef(*100):	-20531	-14906
1st Coef(*100):	270252	139749
2nd Coef(*100):	1917	1284
H2O adj(mg/l*10k):	393	321

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: *Print Test*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
04/28/2023 11:22

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:22
02 Std. Sol.	0.020	11:23
03 Room Air	0.000	11:23
04 Std. Sol.	0.021	11:24
05 Room Air	0.000	11:25
06 Std. Sol.	0.022	11:25
07 Room Air	0.000	11:26

08 Sim Temp = 34.0°C

Simul Ser No = MP5318

Std Sol No = 202201A

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks:

LOW AC Check - 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
04/28/2023 11:27

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:27
02 Std. Sol.	0.300	11:28
03 Room Air	0.000	11:29
04 Std. Sol.	0.303	11:29
05 Room Air	0.000	11:30
06 Std. Sol.	0.303	11:31
07 Room Air	0.000	11:31

08 Sim Temp = 34.0°C

Simul Ser No = MP3067

Std Sol No = 202201F

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks:

High AC Check - 0.300AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
04/28/2023 11:33

Test	AC	Time
01 Room Air	0.000	11:34
02 *Subject Test	INT*	11:34
03 Room Air	0.000	11:35

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 26021080A1
Cyl No = 16
Expiration Date = 10/05/2023
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Interferent check*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
04/28/2023 11:35

Test	AC	Time
01 Diagnostic	OK	11:36
02 Room Air	0.000	11:37
03 *Subject Test	RFI*	11:37
04 Room Air	0.000	11:37

*Invalid Test
 Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 26021080A1
Cyl No = 16
Expiration Date = 10/05/2023
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks:

RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
04/28/2023 11:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:39
02 Std. Gas	0.080	11:39
03 Room Air	0.000	11:40
04 Std. Gas	0.080	11:40
05 Room Air	0.000	11:41
06 Std. Gas	0.080	11:41
07 Room Air	0.000	11:41

Lot No = 26021080A1
Cyl No = 16
Exp Date = 10/05/2023
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry Cal check #1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
04/28/2023 11:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:42
02 Std. Gas	0.080	11:43
03 Room Air	0.000	11:43
04 Std. Gas	0.080	11:44
05 Room Air	0.000	11:44
06 Std. Gas	0.080	11:45
07 Room Air	0.000	11:45

Lot No = 26021080A1
Cyl No = 16
Exp Date = 10/05/2023
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #46

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.14.00 09/16
04/28/2023 11:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:46
02 Std. Gas	0.080	11:46
03 Room Air	0.000	11:47
04 Std. Gas	0.080	11:47
05 Room Air	0.000	11:48
06 Std. Gas	0.080	11:48
07 Room Air	0.000	11:49

Lot No = 26021080A1
Cyl No = 16
Exp Date = 10/05/2023
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry Cal check #7-9

Form 106-I8000