



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 0503 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # 34921080A1 Cyl. # 31)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: MIPD
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
 - Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
 - Display: 507 psi Regulator: 500 psi
 - Display and Regulator ± 50 psi of each other Yes or No
 - Gas tank tare necessary? Yes or No
 - If Yes, display readings after tare (Level 3,M,C,G):
 - Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
 - Wet Calibration Check - Low AC (Level 1,C)
 - Known Value ≤ 0.03 AC: 0.020 AC
 - Sim. Ser #: MP5318
 - Lot #: 202201A
 - Exp. Date: 11Jan24
 - Results ± 0.005 of known AC

Toxicology Unit - Breath Alcohol Section	Qualtrax ID: 11698 Revision: 1
Intoxilyzer 8000 Annual Inspection	Status: Published Published: 03/24/2023
UNCONTROLLED WHEN PRINTED	Page 1 of 2

2. Configure simulator for the following test (Level 1,S).
 Wet Calibration Check - High AC (Level 1,C)
 Known Value \geq 0.25 AC: 0.300 AC
 Sim. Ser #: MP3067
 Lot #: 202201F
 Exp. Date: 18Jun24
 Results \pm 5% AC of known AC
3. Configure dry gas standard for the remaining tests (Level 1,S).
 Known Value : 0.080 AC
 Gas Cylinder Lot #: 26021080A1
 Cylinder #: 16
 Exp. Date: 10/5/23
4. Interferent Check (Level 1,B)
 Known Value: 0.10 AC + 0.05% Acetone
 Sim. Ser #: DR7345
 Lot #: ICS7
 Exp. Date: N/A
 Display reads "Interferent Detect"
5. RFI Check (CMS Mode)
 Display reads "RFI Detect"
6. Dry Calibration Check (Level 1,C)
 Test 1 0.081 Test 4 0.082 Test 7 0.082
 Test 2 0.082 Test 5 0.082 Test 8 0.082
 Test 3 0.081 Test 6 0.083 Test 9 0.083
 Average 0.082
 Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4.

[Signature]
 Inspector Signature

28Apr2023
 Date

[Signature]
 Reviewer

02May2023
 Date

Toxicology Unit - Breath Alcohol Section	Qualtrax ID: 11698 Revision: 1
Intoxilyzer 8000 Annual Inspection	Status: Published Published: 03/24/2023
UNCONTROLLED WHEN PRINTED	Page 2 of 2

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = TOXL 8164.14.00 09/16
04/28/2023 15:56

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Wet
Standard Value? 0.020
Standard Solution #? 202201A
Sim Serial #? MP5318
Oper No? 133237

Flow Cal. Date: 08/20/2015
Slope 663
Intercept -633300

IR Calibration Date: 06/12/2017
 3um 9um

0th Coef(*100): -6741 -15632
1st Coef(*100): 256583 132194
2nd Coef(*100): 3028 1131
H2O adj(mg/l*10k): 354 358

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: Pnnt Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = TOXL 8164.14.00 09/16
04/28/2023 15:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:57
02 Std. Sol.	0.019	15:58
03 Room Air	0.000	15:58
04 Std. Sol.	0.020	15:59
05 Room Air	0.000	16:00
06 Std. Sol.	0.021	16:00
07 Room Air	0.000	16:01

08 Sim Temp = 34.0°C

Simul Ser No = MP5318
Std Sol No = 202201A
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks:

LOW AC CHECK - 0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = TOXL 8164.14.00 09/16
04/28/2023 16:01

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:02
02 Std. Sol.	0.301	16:03
03 Room Air	0.000	16:03
04 Std. Sol.	0.305	16:04
05 Room Air	0.000	16:05
06 Std. Sol.	0.304	16:05
07 Room Air	0.000	16:06

08 Sim Temp = 34.0°C

Simul Ser No = MP3067
Std Sol No = 202201F
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High Ac Check - 0.300Ac

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = TOXL 8164.14.00 09/16
04/28/2023 16:08

Test	AC	Time
01 Room Air	0.000	16:08
02 *Subject Test	INT*	16:09
03 Room Air	0.000	16:09

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 26021080A1
Cyl No = 16
Expiration Date = 10/05/2023
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Interferent check*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = TOXL 8164.14.00 09/16
04/28/2023 16:10

Test	AC	Time
01 Diagnostic	OK	16:10
02 Room Air	RFI*	16:10
03 Room Air	0.000	16:11

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 26021080A1
Cyl No = 16
Expiration Date = 10/05/2023
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = TOXL 8164.14.00 09/16
04/28/2023 16:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:12
02 Std. Gas	0.081	16:12
03 Room Air	0.000	16:13
04 Std. Gas	0.082	16:13
05 Room Air	0.000	16:14
06 Std. Gas	0.081	16:14
07 Room Air	0.000	16:14

Lot No = 26021080A1
Cyl No = 16
Exp Date = 10/05/2023
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry Cal check #1-3

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = TOXL 8164.14.00 09/16
04/28/2023 16:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:15
02 Std. Gas	0.082	16:16
03 Room Air	0.000	16:16
04 Std. Gas	0.082	16:17
05 Room Air	0.000	16:17
06 Std. Gas	0.083	16:17
07 Room Air	0.000	16:18

Lot No = 26021080A1
Cyl No = 16
Exp Date = 10/05/2023
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry Cal Check #4-4

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = TOXL 8164.14.00 09/16
04/28/2023 16:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:19
02 Std. Gas	0.082	16:19
03 Room Air	0.000	16:20
04 Std. Gas	0.082	16:20
05 Room Air	0.000	16:21
06 Std. Gas	0.083	16:21
07 Room Air	0.000	16:21

Lot No = 26021080A1
Cyl No = 16
Exp Date = 10/05/2023
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry Cal check #7-9

Form 106-I8000