



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-003059	Instrument Location New Salem ND.
Reason for Install/Repair	
<input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) Annual Inspection at Crime Laboratory	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): **None**
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date 4-29-2023
---	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>[Signature]</i>	Certified Date 01May2023
---	------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = NS 8164.14.00 09/16
04/29/2023 08:37

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
|
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
|
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer (Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 01923080A3
Standard Cyl #? 24
Standard Expiration? 02/05/2025
Oper No? 119901

Flow Cal. Date: 11/28/2018
Slope 589
Intercept -276335

IR Calibration Date: 12/04/2018
 3um 9um

	3um	9um
0th Coef (*100):	-25330	-15825
1st Coef (*100):	278286	137135
2nd Coef (*100):	1992	1257
H2O adj (mg/l*10k):	525	321

***** Printer Test End *****



Operator Signature
SCOTT AUSTIN

Remarks: *Print Test for Install*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = NS 8164.14.00 09/16
04/29/2023 08:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:40
02 Std. Gas	0.082	08:40
03 Room Air	0.000	08:41
04 Std. Gas	0.081	08:41
05 Room Air	0.000	08:42
06 Std. Gas	0.082	08:42
07 Room Air	0.000	08:43

Lot No = 01923080A3
Cyl No = 24
Exp Date = 02/05/2025
County = 30 Oper No. = 119901



Operator Signature
SCOTT AUSTIN

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = NS 8164.14.00 09/16
04/29/2023 08:46

Test	AC	Time
01 Room Air	0.000	08:51
02 *Subject Test	RFI*	08:51
03 Room Air	0.000	08:51

*Invalid Test
Inhibited - RFI

Sub Name = RFI CHECK, INSTALL NA
Sub DOB = 01/01/2001
Sub Sex = Unknown Weight = NA
Test = OTH Cit = NA
Dr. Lic. = ND/NA
Lot No = 01923080A3
Cyl No = 24
Expiration Date = 02/05/2025
County = 30 Oper No. = 119901



Operator Signature
SCOTT AUSTIN

Remarks:

Form 106-I8000