



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number 80-005958	Instrument Location Oliver County Sheriff's Department
Reason for Install/Repair	
<input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>Annual Inspection at Crime Lab</u>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): Annual Inspection at Crime Lab
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Shannon P. Liddell</i>	Date 04/30/2023
--	--------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Aime Wahl</i>	Certified Date 08May2023
---	-----------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005958  
Location = OLIV                8164.14.00 09/16  
04/30/2023                      17:08

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            34921080A1  
Standard Cyl #?            58  
Standard Expiration?       02/05/2024  
Oper No?                    020904

Flow Cal. Date:            06/21/2013  
Slope                      716  
Intercept                  -760436

IR Calibration Date:       04/07/2015  
                                    3um                      9um

	3um	9um
0th Coef(*100):	-17179	-17694
1st Coef(*100):	259328	130890
2nd Coef(*100):	3473	1654
H2O adj(mg/l*10k):	425	422

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
SHANNON LIEBOLD

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005958  
Location = OLIV      8164.14.00 09/16  
04/30/2023      17:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:19
02 Std. Gas	0.079	17:19
03 Room Air	0.000	17:20
04 Std. Gas	0.080	17:20
05 Room Air	0.000	17:21
06 Std. Gas	0.080	17:21
07 Room Air	0.000	17:22

Lot No = 34921080A1  
Cyl No = 58  
Exp Date = 02/05/2024  
County = 33      Oper No. = 020904



Operator Signature  
SHANNON LIEBOLD

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005958  
Location = OLIV                    8164.14.00 09/16  
04/30/2023                            17:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	17:23
02 Room Air	0.000	17:24

\*Invalid Test  
Inhibited - RFI

Lot No = 34921080A1  
Cyl No = 58  
Exp Date = 02/05/2024  
County = 33                            Oper No. = 020904



Operator Signature  
SHANNON LIEBOLD

Remarks:

Form 106-I8000