



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-607097	Instrument Location Rolla, ND
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>Annual Inspection at Crime Laboratory</u>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): not applicable
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test

Field Inspector Signature <i>Zachary Smith</i> Zachary Smith	Date 5/10/23
--	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>[Signature]</i>	Certified Date 11 May 2023
---	--------------------------------------

#6, C, 6 + 10 were checked off by AEN. These were sent with paperwork.
 #9: AEN confirmed w/ agency that these records are filed at their location.

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = ROLL 8164.16.00 09/18
05/10/2023 15:04

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=|
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=|
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34921080A1
Standard Cyl #? 61
Standard Expiration? 02/05/2024
Oper No? 131486

Flow Cal. Date: 09/03/2020
Slope 660
Intercept -636264

IR Calibration Date: 09/03/2020
3um 9um

0th Coef(*100):	-36810	-25919
1st Coef(*100):	254044	139068
2nd Coef(*100):	1858	1258
H2O adj(mg/l*10k):	903	584

***** Printer Test End *****



Operator Signature
ZACHARY SMITH

Remarks:

Intox install test

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = ROLL 8164.16.00 09/18
05/10/2023 15:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:05
02 Std. Gas	0.080	15:05
03 Room Air	0.000	15:06
04 Std. Gas	0.080	15:06
05 Room Air	0.000	15:07
06 Std. Gas	0.080	15:07
07 Room Air	0.000	15:08

Lot No = 34921080A1
Cyl No = 61
Exp Date = 02/05/2024
County = 40 Oper No. = 131486


Operator Signature
ZACHARY SMITH

Remarks:

Intox Install Test
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = ROLL 8164.16.00 09/18
05/10/2023 15:09

Test	AC	Time
01 Room Air	RFI*	15:10
02 Room Air	0.000	15:11

*Invalid Test
Inhibited - RFI

Sub Name = RFI TEST, RFI RFI
Sub DOB = 04/24/1993
Sub Sex = Male Weight = 123
Test = OTH Cit = NA
Dr. Lic. = ND/NA
Lot No = 34921080A1
Cyl No = 61
Expiration Date = 02/05/2024
County = 40 Oper No. = 131486



Operator Signature
ZACHARY SMITH

Remarks: Test
Install Install Intox₂

Form 106-I8000