



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6667 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # 26021080A1 Cyl. # 28)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: GLDN
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
Display: 980 psi Regulator: 1000 psi
Display and Regulator ± 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP5318
Lot #: 202201A
Exp. Date: 11Jan24
 Results ± 0.005 of known AC

Toxicology Unit - Breath Alcohol Section	Qualtrax ID: 11698 Revision: 1
Intoxilyzer 8000 Annual Inspection	Status: Published Published: 03/24/2023
UNCONTROLLED WHEN PRINTED	Page 1 of 2

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = TOXL 8164.14.00 09/16
05/11/2023 12:59

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%\$^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%\$^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 26021080A1
Standard Cyl #? 28
Standard Expiration? 10/05/2023
Oper No? 133237

Flow Cal. Date: 05/22/2020
Slope 678
Intercept -711846

IR Calibration Date: 05/22/2020
 3um 9um

0th Coef(*100): -22305 -20910
1st Coef(*100): 281966 134470
2nd Coef(*100): 1570 1173
H2O adj(mg/l*10k): 336 351

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: Pnht Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = TOXL 8164.14.00 09/16
05/11/2023 13:00

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:00
02 Std. Sol.	0.020	13:01
03 Room Air	0.000	13:02
04 Std. Sol.	0.021	13:03
05 Room Air	0.000	13:03
06 Std. Sol.	0.021	13:04
07 Room Air	0.000	13:04

08 Sim Temp = 34.0°C

Simul Ser No = MP5318
Std Sol No = 202201A
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Low AC Check - 0.0201C

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = TOXL 8164.14.00 09/16
05/11/2023 13:10

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:11
02 Std. Sol.	0.306	13:12
03 Room Air	0.000	13:12
04 Std. Sol.	0.306	13:13
05 Room Air	0.000	13:14
06 Std. Sol.	0.305	13:14
07 Room Air	0.000	13:15

08 Sim Temp = 34.0°C

Simul Ser No = MP3067

Std Sol No = 202201F

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High Ac Check - 0.300Ac

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = TOXL 8164.14.00 09/16
05/11/2023 13:16

Test	AC	Time
01 Room Air	0.000	13:17
02 *Subject Test	INT*	13:17
03 Room Air	0.000	13:18

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 34921080A1
Cyl No = 33
Expiration Date = 02/05/2024
County = 08 Oper No. = 133237


Operator Signature
ANNA NAREHOOD

Remarks: *Interferent check*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = TOXL 8164.14.00 09/16
05/11/2023 13:19

Test	AC	Time
01 Diagnostic	OK	13:19
02 Room Air	0.000	13:20
03 *Subject Test	RFI*	13:20
04 Room Air	0.000	13:20

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 34921080A1
Cyl No = 33
Expiration Date = 02/05/2024
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = TOXL 8164.14.00 09/16
05/11/2023 13:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:22
02 Std. Gas	0.081	13:22
03 Room Air	0.000	13:23
04 Std. Gas	0.081	13:23
05 Room Air	0.000	13:24
06 Std. Gas	0.081	13:24
07 Room Air	0.000	13:25

Lot No = 34921080A1
Cyl No = 33
Exp Date = 02/05/2024
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #1-3

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = TOXL 8164.14.00 09/16
05/11/2023 13:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:26
02 Std. Gas	0.081	13:26
03 Room Air	0.000	13:27
04 Std. Gas	0.081	13:27
05 Room Air	0.000	13:27
06 Std. Gas	0.081	13:28
07 Room Air	0.000	13:28

Lot No = 34921080A1
Cyl No = 33
Exp Date = 02/05/2024
County = 08 Oper No. = 133237

Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = TOXL 8164.14.00 09/16
05/11/2023 13:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:29
02 Std. Gas	0.081	13:30
03 Room Air	0.000	13:30
04 Std. Gas	0.081	13:30
05 Room Air	0.000	13:31
06 Std. Gas	0.081	13:31
07 Room Air	0.000	13:32

Lot No = 34921080A1
Cyl No = 33
Exp Date = 02/05/2024
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #7-9

Form 106-I8000