



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 5951 Inspection Location: ToxL

A. Pre-Inspection

1. Items with Instrument:
  - Gas Cylinder Yes or No (If Yes, Lot # 26021080A1 Cyl. # 29)
  - Keys Yes or No
  - Power Cord Yes or No
2.  Download Data
3.  Upload Operator File
4.  Current Location Code: LANG
5.  Battery Check
  - Was the external battery pack replaced? Yes or No
6.  O-Rings
  - Replaced Simulator O-Ring Yes or No
  - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1.  Diagnostics passed and instrument in "Ready" mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G).  
Display: 876 psi Regulator: 900 psi  
Display and Regulator ± 50 psi of each other Yes or No  
Gas tank tare necessary? Yes or No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1.  Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value ≤ 0.03 AC: 0.020 AC  
Sim. Ser #: MP5318  
Lot #: 202201A  
Exp. Date: 11Jan24  
 Results ± 0.005 of known AC

Toxicology Unit - Breath Alcohol Section	Qualtrax ID: 11698 Revision: 1
Intoxilyzer 8000 Annual Inspection	Status: Published Published: 03/24/2023
UNCONTROLLED WHEN PRINTED	Page 1 of 2





Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005951  
Location = TOXL      8164.14.00 09/16  
05/11/2023      15:05

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:06
02 Std. Sol.	0.020	15:07
03 Room Air	0.000	15:07
04 Std. Sol.	0.020	15:08
05 Room Air	0.000	15:09
06 Std. Sol.	0.020	15:09
07 Room Air	0.000	15:10

08 Sim Temp = 34.0°C

Simul Ser No = MP5318  
Std Sol No = 202201A  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: *LOW AC Check- 0.020AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005951  
Location = TOXL      8164.14.00 09/16  
05/11/2023      15:16

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:17
02 Std. Sol.	0.304	15:17
03 Room Air	0.000	15:18
04 Std. Sol.	0.305	15:19
05 Room Air	0.000	15:19
06 Std. Sol.	0.305	15:20
07 Room Air	0.000	15:20

08 Sim Temp = 34.0°C

Simul Ser No = MP3067  
Std Sol No = 202201F  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: High AC Check - 0.300AC

Form 106-I8000






Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005951  
Location = TOXL      8164.14.00 09/16  
05/11/2023      15:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:27
02 Std. Gas	0.081	15:27
03 Room Air	0.000	15:28
04 Std. Gas	0.082	15:28
05 Room Air	0.000	15:29
06 Std. Gas	0.081	15:29
07 Room Air	0.000	15:30

Lot No = 34921080A1  
Cyl No = 33  
Exp Date = 02/05/2024  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #1-3

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005951  
Location = TOXL      8164.14.00 09/16  
05/11/2023      15:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:31
02 Std. Gas	0.082	15:31
03 Room Air	0.000	15:32
04 Std. Gas	0.082	15:32
05 Room Air	0.000	15:33
06 Std. Gas	0.081	15:33
07 Room Air	0.000	15:34

Lot No = 34921080A1  
Cyl No = 33  
Exp Date = 02/05/2024  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA NAREHOOD

Remarks: Dry Cal Check #4-6

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005951  
Location = TOXL      8164.14.00 09/16  
05/11/2023      15:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:35
02 Std. Gas	0.082	15:35
03 Room Air	0.000	15:36
04 Std. Gas	0.082	15:36
05 Room Air	0.000	15:37
06 Std. Gas	0.081	15:37
07 Room Air	0.000	15:38

Lot No = 34921080A1  
Cyl No = 33  
Exp Date = 02/05/2024  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: *Dry cal check #7-9*

Form 106-I8000