



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 7091 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # 26021080A1 Cyl. # 9)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: MOTT
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
Display: 987 psi Regulator: 1000 psi
Display and Regulator ± 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP5318
Lot #: 202201A
Exp. Date: 11 Jan 24
 Results ± 0.005 of known AC

Toxicology Unit - Breath Alcohol Section	Qualtrax ID: 11698 Revision: 1
Intoxilyzer 8000 Annual Inspection	Status: Published Published: 03/24/2023
UNCONTROLLED WHEN PRINTED	Page 1 of 2

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007091
Location = TOXL 8164.16.00 09/18
05/12/2023 14:20

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 26021080A1
Standard Cyl #? 9
Standard Expiration? 10/05/2023
Oper No? 133237

Flow Cal. Date: 09/25/2018
Slope 681
Intercept -651030

IR Calibration Date: 09/25/2018
 3um 9um

0th Coef(*100): -39121 -20906
1st Coef(*100): 240667 132345
2nd Coef(*100): 3960 1516
H2O adj(mg/l*10k): 899 428

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: *Print Test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007091
Location = TOXL 8164.16.00 09/18
05/12/2023 14:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:21
02 Std. Sol.	0.018	14:22
03 Room Air	0.000	14:22
04 Std. Sol.	0.018	14:23
05 Room Air	0.000	14:23
06 Std. Sol.	0.018	14:24
07 Room Air	0.000	14:25

08 Sim Temp = 34.0°C

Simul Ser No = MP5318

Std Sol No = 202201A

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks:

Low Ac Check- 0.020Ac

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007091
Location = TOXL 8164.16.00 09/18
05/12/2023 14:25

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:26
02 Std. Sol.	0.298	14:27
03 Room Air	0.000	14:27
04 Std. Sol.	0.302	14:28
05 Room Air	0.000	14:28
06 Std. Sol.	0.302	14:29
07 Room Air	0.000	14:30

08 Sim Temp = 34.0°C

Simul Ser No = MP3067
Std Sol No = 202201F
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High Ae check- 0.300 Ae

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007091
Location = TOXL 8164.16.00 09/18
05/12/2023 14:31

Test	AC	Time
01 Room Air	0.000	14:31
02 *Subject Test	INT*	14:32
03 Room Air	0.000	14:33

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 34921080A1
Cyl No = 49
Expiration Date = 02/05/2024
County = 08 Oper No. = 133237

Operator Signature
ANNA NAREHOOD

Remarks: *Interferent check*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007091
Location = TOXL 8164.16.00 09/18
05/12/2023 14:33

Test	AC	Time
01 Diagnostic	OK	14:34
02 Room Air	0.000	14:34
03 *Subject Test	RFI*	14:34
04 Room Air	0.000	14:35

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 34921080A1
Cyl No = 49
Expiration Date = 02/05/2024
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007091
Location = TOXL 8164.16.00 09/18
05/12/2023 14:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:36
02 Std. Gas	0.078	14:36
03 Room Air	0.000	14:36
04 Std. Gas	0.078	14:37
05 Room Air	0.000	14:37
06 Std. Gas	0.078	14:38
07 Room Air	0.000	14:38

Lot No = 34921080A1
Cyl No = 49
Exp Date = 02/05/2024
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Drycal check #1-3*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007091
Location = TOXL 8164.16.00 09/18
05/12/2023 14:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:39
02 Std. Gas	0.078	14:39
03 Room Air	0.000	14:40
04 Std. Gas	0.079	14:40
05 Room Air	0.000	14:41
06 Std. Gas	0.078	14:41
07 Room Air	0.000	14:42

Lot No = 34921080A1
Cyl No = 49
Exp Date = 02/05/2024
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Dry cal check # 4-6*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007091
Location = TOXL 8164.16.00 09/18
05/12/2023 14:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:42
02 Std. Gas	0.079	14:43
03 Room Air	0.000	14:43
04 Std. Gas	0.079	14:44
05 Room Air	0.000	14:44
06 Std. Gas	0.079	14:45
07 Room Air	0.000	14:45

Lot No = 34921080A1
Cyl No = 49
Exp Date = 02/05/2024
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Dry cal check #7-9*

Form 106-I8000