




INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

| | |
|---|---|
| Serial Number 80-006673 | Instrument Location Burke County Sheriff's Shop |
| Reason for Install/Repair | |
| <input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) Annual Inspection at Crime Laboratory | |

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): **N/A**
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

| | | |
|---|-------|--------------------------|
| Field Inspector Signature  | #4172 | Date 5/12/2023 |
|---|-------|--------------------------|

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

| | |
|--|------------------------------------|
| Reviewed/Certified By  | Certified Date 16May2023 |
|--|------------------------------------|

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006673
Location = BURK 8164.14.00 09/16
05/12/2023 17:53

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34921080A1
Standard Cyl #? 54
Standard Expiration? 02/05/2024
Oper No? 132664

Flow Cal. Date: 03/30/2017
Slope 697
Intercept -667725

IR Calibration Date: 03/30/2017
 3um 9um

| | | |
|--------------------|--------|--------|
| 0th Coef(*100): | -9136 | -17613 |
| 1st Coef(*100): | 267126 | 133333 |
| 2nd Coef(*100): | 5438 | 1698 |
| H2O adj(mg/l*10k): | 684 | 589 |

***** Printer Test End *****

 #4172
Operator Signature
CHRISTOPHER ALMLIE

Remarks: good test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006673
Location = BURK 8164.14.00 09/16
05/12/2023 17:56

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 17:57 |
| 02 Std. Gas | 0.081 | 17:57 |
| 03 Room Air | 0.000 | 17:58 |
| 04 Std. Gas | 0.080 | 17:58 |
| 05 Room Air | 0.000 | 17:59 |
| 06 Std. Gas | 0.080 | 17:59 |
| 07 Room Air | 0.000 | 17:59 |

Lot No = 34921080A1
Cyl No = 54
Exp Date = 02/05/2024
County = 07 Oper No. = 132664

 #4172
Operator Signature
CHRISTOPHER ALMLIE

Remarks: good test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006673
Location = BURK 8164.14.00 09/16
05/12/2023 18:00

| Test | AC | Time |
|-------------|------|-------|
| 01 Room Air | RFI* | 18:02 |
| 02 Room Air | RFI* | 18:02 |

*Invalid Test
 Inhibited - RFI

Sub Name = TEST, TEST TEST
Sub DOB = 05/12/2000
Sub Sex = Male Weight = NA
Test = OTH Cit = NA
Dr. Lic. = ND/NA
Lot No = 34921080A1
Cyl No = 54
Expiration Date = 02/05/2024
County = 07 Oper No. = 132664

 #4172
Operator Signature
CHRISTOPHER ALMLIE

Remarks: *good test*

Form 106-I8000